

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TULL01
Lab Code: CHEM Case No.: Q1012 SAS No.: Q1012 SDG NO.: Q1012
Lab File ID: VN085390.D Date Analyzed: 01/07/2025
Instrument ID: MSVOA_N Time Analyzed: 10:30
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	28915	7.81	149310	9.09	139056	11.86
UPPER LIMIT	57830	8.306	298620	9.594	278112	12.359
LOWER LIMIT	14457.5	7.306	74655	8.594	69528	11.359
EPA SAMPLE NO.						
001 WILLETS PT BLVD (DEC)	22808	7.81	111854	9.10	108816	11.86
VN0107WBL01	26828	7.81	128912	9.09	114923	11.86
VN0107WBS01	29795	7.81	157585	9.09	144288	11.86
VN0107WBSD01	31547	7.81	161684	9.09	151196	11.86

IS1 = Bromochloromethane
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.