



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

### Initial and Continuing Calibration Blank Summary

**Client:** Tully Environmental, Inc

**SDG No.:** Q1013

**Project:** Transfer Station-SPDES

**RunNo.:** LB134167

| Analyte                                | Units | Result   | Acceptance<br>Limits | Conc<br>Qual | MDL   | RDL | Analysis<br>Date |
|----------------------------------------|-------|----------|----------------------|--------------|-------|-----|------------------|
| Sample ID: <b>ICB1</b><br>Ammonia as N | mg/L  | < 0.0500 | 0.0500               | U            | 0.045 | 0.1 | 01/06/2025       |
| Sample ID: <b>CCB1</b><br>Ammonia as N | mg/L  | < 0.0500 | 0.0500               | U            | 0.045 | 0.1 | 01/06/2025       |
| Sample ID: <b>CCB2</b><br>Ammonia as N | mg/L  | < 0.0500 | 0.0500               | U            | 0.045 | 0.1 | 01/06/2025       |
| Sample ID: <b>CCB3</b><br>Ammonia as N | mg/L  | < 0.0500 | 0.0500               | U            | 0.045 | 0.1 | 01/06/2025       |

### Initial and Continuing Calibration Blank Summary

**Client:** Tully Environmental, Inc

**SDG No.:** Q1013

**Project:** Transfer Station-SPDES

**RunNo.:** LB134167

| Analyte | Units | Result | Acceptance<br>Limits | Conc<br>Qual | MDL | RDL | Analysis<br>Date |
|---------|-------|--------|----------------------|--------------|-----|-----|------------------|
|---------|-------|--------|----------------------|--------------|-----|-----|------------------|