

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1013
Project:	Transfer Station-SPDES	Sample ID:	LB134202BS
Client ID:	LB134202BSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Oil and Grease	mg/L	+/-18	16.7		16.9		1	1.19		01/09/2025

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1013
Project:	Transfer Station-SPDES	Sample ID:	P5386-05
Client ID:	MOO-24-00397DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	0.057	J	0.054	J	1	5		01/06/2025

Duplicate Sample Summary

Client: Tully Environmental, Inc	SDG No.: Q1013
Project: Transfer Station-SPDES	Sample ID: P5386-05
Client ID: MOO-24-00397MSD	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/L	+/-20	1.00		1.10		1	10		01/06/2025

Duplicate Sample Summary

Client: Tully Environmental, Inc	SDG No.: Q1013
Project: Transfer Station-SPDES	Sample ID: Q1012-04
Client ID: 002 35TH AVE (DEC)DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	35.5		36.8		1	3.6		01/07/2025

Duplicate Sample Summary

Client:	Tully Environmental, Inc	SDG No.:	Q1013
Project:	Transfer Station-SPDES	Sample ID:	Q1013-01
Client ID:	001 WILLETS PT BLVD (JAN)DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
BOD5	mg/L	+/-20	85.2		81.8		1	4.13		01/08/2025