

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1013

Project: Transfer Station-SPDES

Run No.: LB134181

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134181BS							
TSS	mg/L	550	542		98	1	90-110	01/07/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1013

Project: Transfer Station-SPDES

Run No.: LB134191

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134191BS							
BOD5	mg/L	198	200		101	1	84.6-115.4	01/08/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1013

Project: Transfer Station-SPDES

Run No.: LB134202

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134202BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	01/09/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1013

Project: Transfer Station-SPDES

Run No.: LB134202

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134202BSD							
Oil and Grease	mg/L	20.0	16.9		84	1	78-114	01/09/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q1013

Project: Transfer Station-SPDES

Run No.: LB134167

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165937BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	01/06/2025