

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134201

Review By	kareem	Review On	1/9/2025 3:48:43 PM
Supervise By	Jaswal	Supervise On	1/20/2025 7:26:28 AM
STD. NAME	STD REF.#		
ICAL Standard	mp83935 mp83936 mp83937 mp83938 mp83939 mp83941		
ICV Standard	mp83942		
CCV Standard	mp83945		
ICSA Standard	mp83943 mp83944		
CRI Standard	mp83941		
LCS Standard			
Chk Standard	mp83948 mp83949		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	01/08/25 10:53		Kareem	OK
2	S1	S1	CAL2	01/08/25 10:57		Kareem	OK
3	S2	S2	CAL3	01/08/25 11:02		Kareem	OK
4	S3	S3	CAL4	01/08/25 11:06		Kareem	OK
5	S4	S4	CAL5	01/08/25 11:10		Kareem	OK
6	S5	S5	CAL6	01/08/25 11:15		Kareem	OK
7	ICV01	ICV01	ICV	01/08/25 11:19	ICV Fail for Be,Ca,Mg,Mn,Se,Vanadium	Kareem	OK
8	LLICV01	LLICV01	LLICV	01/08/25 11:30		Kareem	OK
9	ICB01	ICB01	ICB	01/08/25 11:34		Kareem	OK
10	CRI01	CRI01	CRDL	01/08/25 11:38		Kareem	OK
11	ICSA01	ICSA01	ICSA	01/08/25 11:43		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	01/08/25 11:47		Kareem	OK
13	ICSADL	ICSADL	ICSA	01/08/25 11:51		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	01/08/25 11:55		Kareem	OK
15	CCV01	CCV01	CCV	01/08/25 12:00		Kareem	OK
16	CCB01	CCB01	CCB	01/08/25 12:04		Kareem	OK
17	PB165970BL	PB165970BL	MB	01/08/25 12:08		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134201

Review By	kareem	Review On	1/9/2025 3:48:43 PM
Supervise By	Jaswal	Supervise On	1/20/2025 7:26:28 AM

STD. NAME	STD REF.#
ICAL Standard	mp83935 mp83936 mp83937 mp83938 mp83939 mp83941
ICV Standard	mp83942
CCV Standard	mp83945
ICSA Standard	mp83943 mp83944
CRI Standard	mp83941
LCS Standard	
Chk Standard	mp83948 mp83949

18	PB165970BS	PB165970BS	LCS	01/08/25 12:13	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
19	Q1013-01	001 WILLETS PT BLV	SAM	01/08/25 12:17		Kareem	OK
20	Q1013-02	002 35TH AVE (JAN)	SAM	01/08/25 12:21		Kareem	OK
21	Q1013-02DUP	002 35TH AVE (JAN)	DUP	01/08/25 12:25		Kareem	OK
22	Q1013-02L	002 35TH AVE (JAN)	SD	01/08/25 12:30		Kareem	OK
23	Q1013-02MS	002 35TH AVE (JAN)	MS	01/08/25 12:34	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
24	Q1013-02MSD	002 35TH AVE (JAN)	MSD	01/08/25 12:38	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
25	Q1013-02A	002 35TH AVE (JAN)	PS	01/08/25 12:42	0.1ml of m6010 and m6001 were added to 10ml of the sample	Kareem	OK
26	CCV02	CCV02	CCV	01/08/25 12:46		Kareem	OK
27	CCB02	CCB02	CCB	01/08/25 12:50		Kareem	OK
28	PB165979BL	PB165979BL	MB	01/08/25 12:55		Kareem	OK
29	PB165979BS	PB165979BS	LCS	01/08/25 12:59		Kareem	OK
30	Q1025-01	GREENBROOK-COM	SAM	01/08/25 13:03		Kareem	OK
31	Q1027-01	EO-03-010725	SAM	01/08/25 13:07		Kareem	OK
32	Q1028-01	OR-03-010725	SAM	01/08/25 13:11		Kareem	OK
33	Q1029-03	HR-01-01072025	SAM	01/08/25 13:15	MS & MSD fail for 50% parameter	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB134201

Review By	kareem	Review On	1/9/2025 3:48:43 PM
Supervise By	Jaswal	Supervise On	1/20/2025 7:26:28 AM

STD. NAME	STD REF.#
ICAL Standard	mp83935 mp83936 mp83937 mp83938 mp83939 mp83941
ICV Standard	mp83942
CCV Standard	mp83945
ICSA Standard	mp83943 mp83944
CRI Standard	mp83941
LCS Standard	
Chk Standard	mp83948 mp83949

34	Q1029-03DUP	HR-01-01072025DUP	DUP	01/08/25 13:20	MS & MSD fail for 50% parameter	Kareem	Not Ok
35	Q1029-03L	HR-01-01072025L	SD	01/08/25 13:24	MS & MSD fail for 50% parameter	Kareem	Not Ok
36	Q1029-03MS	HR-01-01072025MS	MS	01/08/25 13:28	MS & MSD fail for 50% parameter	Kareem	Not Ok
37	Q1029-03MSD	HR-01-01072025MSD	MSD	01/08/25 13:32	MS & MSD fail for 50% parameter	Kareem	Not Ok
38	CCV03	CCV03	CCV	01/08/25 13:36		Kareem	OK
39	CCB03	CCB03	CCB	01/08/25 13:40		Kareem	OK
40	Q1029-03A	HR-01-01072025A	PS	01/08/25 13:44	MS & MSD fail for 50% parameter	Kareem	Not Ok
41	Q1029-01	HR-02-01072025	SAM	01/08/25 13:48		Kareem	OK
42	LR1	LR1	HIGH STD	01/08/25 13:55		Kareem	OK
43	LR2	LR2	HIGH STD	01/08/25 14:00		Kareem	OK
44	CCV04	CCV04	CCV	01/08/25 14:04		Kareem	OK
45	CCB04	CCB04	CCB	01/08/25 14:08		Kareem	OK