



- 2a -

Client:	<u>Tully Environmental, Inc</u>	SDG No.:	<u>Q1013</u>	
Contract:	<u>TULL01</u>	Lab Code:	<u>CHEM</u>	Case No.: <u>Q1013</u> SAS No.: <u>Q1013</u>
Initial Calibration Source:	<u>EPA</u>			
Continuing Calibration Source:	Inorganic Ventures			

Sample ID	Analyte	Result	True Value	%	Acceptance	M	Analysis Date	Analysis Time	Run Number
		ug/L		Recovery	Window (%R)				
LLICV01	Arsenic	19.3	20.0	96	80 - 120	P	01/08/2025	11:30	LB134201
	Cadmium	5.76	6.0	96	80 - 120	P	01/08/2025	11:30	LB134201
	Chromium	9.68	10.0	97	80 - 120	P	01/08/2025	11:30	LB134201
	Copper	21.3	20.0	106	80 - 120	P	01/08/2025	11:30	LB134201
	Iron	93.8	100	94	80 - 120	P	01/08/2025	11:30	LB134201
	Lead	10.6	12.0	89	80 - 120	P	01/08/2025	11:30	LB134201
	Selenium	18.6	20.0	93	80 - 120	P	01/08/2025	11:30	LB134201

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc **SDG No.:** Q1013
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1013 **SAS No.:** Q1013
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	4700	5000	94	90 - 110	P	01/08/2025	12:00	LB134201
	Cadmium	2370	2500	95	90 - 110	P	01/08/2025	12:00	LB134201
	Chromium	999	1000	100	90 - 110	P	01/08/2025	12:00	LB134201
	Copper	1200	1250	96	90 - 110	P	01/08/2025	12:00	LB134201
	Iron	4760	5000	95	90 - 110	P	01/08/2025	12:00	LB134201
	Lead	4760	5000	95	90 - 110	P	01/08/2025	12:00	LB134201
	Selenium	4700	5000	94	90 - 110	P	01/08/2025	12:00	LB134201
CCV02	Arsenic	4970	5000	99	90 - 110	P	01/08/2025	12:46	LB134201
	Cadmium	2440	2500	98	90 - 110	P	01/08/2025	12:46	LB134201
	Chromium	990	1000	99	90 - 110	P	01/08/2025	12:46	LB134201
	Copper	1250	1250	100	90 - 110	P	01/08/2025	12:46	LB134201
	Iron	4660	5000	93	90 - 110	P	01/08/2025	12:46	LB134201
	Lead	4860	5000	97	90 - 110	P	01/08/2025	12:46	LB134201
	Selenium	5010	5000	100	90 - 110	P	01/08/2025	12:46	LB134201
CCV03	Arsenic	4870	5000	97	90 - 110	P	01/08/2025	13:36	LB134201
	Cadmium	2410	2500	96	90 - 110	P	01/08/2025	13:36	LB134201
	Chromium	1030	1000	103	90 - 110	P	01/08/2025	13:36	LB134201
	Copper	1230	1250	99	90 - 110	P	01/08/2025	13:36	LB134201
	Iron	4930	5000	99	90 - 110	P	01/08/2025	13:36	LB134201
	Lead	4830	5000	97	90 - 110	P	01/08/2025	13:36	LB134201
	Selenium	4900	5000	98	90 - 110	P	01/08/2025	13:36	LB134201
CCV04	Arsenic	5070	5000	101	90 - 110	P	01/08/2025	14:04	LB134201
	Cadmium	2410	2500	96	90 - 110	P	01/08/2025	14:04	LB134201
	Chromium	978	1000	98	90 - 110	P	01/08/2025	14:04	LB134201
	Copper	1280	1250	102	90 - 110	P	01/08/2025	14:04	LB134201
	Iron	4940	5000	99	90 - 110	P	01/08/2025	14:04	LB134201
	Lead	4830	5000	96	90 - 110	P	01/08/2025	14:04	LB134201
	Selenium	5180	5000	104	90 - 110	P	01/08/2025	14:04	LB134201