



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc **SDG No.:** Q1013
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1013 **SAS No.:** Q1013
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Arsenic	18.5	20.0	92	40 - 160	P	01/08/2025	11:38	LB134201
	Cadmium	5.74	6.0	96	40 - 160	P	01/08/2025	11:38	LB134201
	Chromium	9.96	10.0	100	40 - 160	P	01/08/2025	11:38	LB134201
	Copper	21.1	20.0	106	40 - 160	P	01/08/2025	11:38	LB134201
	Iron	94.4	100	94	40 - 160	P	01/08/2025	11:38	LB134201
	Lead	10.6	12.0	88	40 - 160	P	01/08/2025	11:38	LB134201
	Selenium	18.0	20.0	90	40 - 160	P	01/08/2025	11:38	LB134201