

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc **SDG No.:** Q1013
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1013 **SAS No.:** Q1013
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	2.49			-20	20	01/08/2025	11:43	LB134201
	Cadmium	0.11	1.0	10	-5	7	01/08/2025	11:43	LB134201
	Chromium	59.7	52.0	115	42	62	01/08/2025	11:43	LB134201
	Copper	18.5	2.0	925	-18	22	01/08/2025	11:43	LB134201
	Iron	102000	101000	101	85600	116500	01/08/2025	11:43	LB134201
	Lead	7.08			-12	12	01/08/2025	11:43	LB134201
	Selenium	-18.6			-20	20	01/08/2025	11:43	LB134201
ICSAB01	Arsenic	106	104	102	88.4	120	01/08/2025	11:47	LB134201
	Cadmium	978	972	101	826	1120	01/08/2025	11:47	LB134201
	Chromium	583	542	108	460	624	01/08/2025	11:47	LB134201
	Copper	492	511	96	434	588	01/08/2025	11:47	LB134201
	Iron	99700	99300	100	84400	114500	01/08/2025	11:47	LB134201
	Lead	55.6	49.0	114	37	61	01/08/2025	11:47	LB134201
	Selenium	26.4	46.0	57	26	66	01/08/2025	11:47	LB134201
ICSA	Arsenic	-72.6			-20	20	01/08/2025	11:51	LB134201
	Cadmium	-5.84	1.0	584	-5	7	01/08/2025	11:51	LB134201
	Chromium	56.7	52.0	109	42	62	01/08/2025	11:51	LB134201
	Copper	12.3	2.0	615	-18	22	01/08/2025	11:51	LB134201
	Iron	99300	101000	98	85600	116500	01/08/2025	11:51	LB134201
	Lead	-12.6			-12	12	01/08/2025	11:51	LB134201
	Selenium	-52.9			-20	20	01/08/2025	11:51	LB134201
ICSAB	Arsenic	66.3	104	64	88.4	120	01/08/2025	11:55	LB134201
	Cadmium	1030	972	106	826	1120	01/08/2025	11:55	LB134201
	Chromium	611	542	113	460	624	01/08/2025	11:55	LB134201
	Copper	522	511	102	434	588	01/08/2025	11:55	LB134201
	Iron	100000	99300	101	84400	114500	01/08/2025	11:55	LB134201
	Lead	36.6	49.0	75	37	61	01/08/2025	11:55	LB134201
	Selenium	18.1	46.0	39	26	66	01/08/2025	11:55	LB134201