

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tully Environmental, Inc **level:** low **sdg no.:** Q1013
contract: TULL01 **lab code:** CHEM **case no.:** Q1013 **sas no.:** Q1013
matrix: Water **sample id:** Q1013-02 **client id:** 002 35TH AVE (JAN)MS
Percent Solids for Sample: NA **Spiked ID:** Q1013-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	392		10.0	U	400	98		P
Cadmium	ug/L	75 - 125	91.2		3.00	U	100	91		P
Chromium	ug/L	75 - 125	187		3.45	J	200	92		P
Copper	ug/L	75 - 125	148		12.4		150	91		P
Iron	ug/L	75 - 125	6050		5040		1500	67	N	P
Lead	ug/L	75 - 125	472		14.1		500	92		P
Selenium	ug/L	75 - 125	401		10.0	U	1000	40	N	P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Tully Environmental, Inc **level:** low **sdg no.:** Q1013
contract: TULL01 **lab code:** CHEM **case no.:** Q1013 **sas no.:** Q1013
matrix: Water **sample id:** Q1013-02 **client id:** 002 35TH AVE (JAN)MSD
Percent Solids for Sample: NA **Spiked ID:** Q1013-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Arsenic	ug/L	75 - 125	393		10.0	U	400	98		P
Cadmium	ug/L	75 - 125	91.1		3.00	U	100	91		P
Chromium	ug/L	75 - 125	195		3.45	J	200	96		P
Copper	ug/L	75 - 125	149		12.4		150	91		P
Iron	ug/L	75 - 125	6310		5040		1500	84		P
Lead	ug/L	75 - 125	469		14.1		500	91		P
Selenium	ug/L	75 - 125	404		10.0	U	1000	40	N	P