

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tully Environmental, Inc **Level:** LOW **SDG No.:** Q1013
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1013 **SAS No.:** Q1013
Matrix: Water **Sample ID:** Q1013-02 **Client ID:** 002 35TH AVE (JAN)DUP
Percent Solids for Sample: NA **Duplicate ID** Q1013-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	10.0	U	10.0	U			P
Cadmium	ug/L	20	3.00	U	3.00	U			P
Chromium	ug/L	20	3.45	J	4.02	J	15		P
Copper	ug/L	20	12.4		12.2		2		P
Iron	ug/L	20	5040		5130		2		P
Lead	ug/L	20	14.1		14.4		2		P
Selenium	ug/L	20	10.0	U	10.0	U			P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tully Environmental, Inc **Level:** LOW **SDG No.:** Q1013
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1013 **SAS No.:** Q1013
Matrix: Water **Sample ID:** Q1013-02MS **Client ID:** 002 35TH AVE (JAN)MSD
Percent Solids for Sample: NA **Duplicate ID** Q1013-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Arsenic	ug/L	20	392		393		0		P
Cadmium	ug/L	20	91.2		91.1		0		P
Chromium	ug/L	20	187		195		4		P
Copper	ug/L	20	148		149		1		P
Iron	ug/L	20	6050		6310		4		P
Lead	ug/L	20	472		469		1		P
Selenium	ug/L	20	401		404		1		P