

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc **SDG No.:** Q1013
Contract: TULL01 **Lab Code:** CHEM **Case No.:** Q1013 **SAS No.:** Q1013

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165970BS							
Arsenic	ug/L	400	387		97	85 - 115	P
Cadmium	ug/L	100	97.6		98	85 - 115	P
Chromium	ug/L	200	203		102	85 - 115	P
Copper	ug/L	150	148		99	85 - 115	P
Iron	ug/L	1500	1380		92	85 - 115	P
Lead	ug/L	500	481		96	85 - 115	P
Selenium	ug/L	1000	939		94	85 - 115	P