

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TULL01

Lab Code: CHEM Case No.: Q1013 SAS No.: Q1013 SDG No.: Q1013

Instrument ID: MSVOA\_N Calibration Date/Time: 01/07/2025 10:30

Lab File ID: VN085390.D Init. Calib. Date(s): 12/20/2024 12/20/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:05 11:42

GC Column: RXI-624 ID: 0.25 (mm)

| COMPOUND              | RRF   | RRF020 | MIN<br>RRF | %D    | MAX%D |
|-----------------------|-------|--------|------------|-------|-------|
| Benzene               | 1.646 | 1.717  | 0.5        | 4.31  |       |
| Toluene               | 1.843 | 1.936  | 0.4        | 5.05  |       |
| Ethyl Benzene         | 1.939 | 1.912  | 0.1        | -1.39 |       |
| m/p-Xylenes           | 0.728 | 0.769  | 0.3        | 5.63  |       |
| o-Xylene              | 0.693 | 0.714  | 0.3        | 3.03  |       |
| 1,2-Dichloroethane-d4 | 2.639 | 2.621  | 0.01       | -0.68 |       |
| Toluene-d8            | 1.465 | 1.505  | 0.01       | 2.73  |       |
| 4-Bromofluorobenzene  | 0.486 | 0.490  | 0.2        | 0.82  |       |

All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.