

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1022 SAS No.: Q1022 SDG NO.: Q1022

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 01/07/2025

Lab File ID: BN035897.D Time Analyzed: 15:03

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1335	7.832	2523	10.62	1187	14.46
UPPER LIMIT	2670	8.332	5046	11.122	2374	14.958
LOWER LIMIT	667.5	7.332	1261.5	10.122	593.5	13.958
EPA SAMPLE NO.						
01 PB165971BL	1101	7.83	2139	10.62	1003	14.46
02 BP-VPB-190A-GW-838-840	1117	7.83	2226	10.62	1401	14.46
03 PB165971BS	1718	7.83	3309	10.62	1504	14.46
04 PB165971BSD	1875	7.83	3596	10.62	1733	14.46

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.