

Surrogate Summary

SDG No.: **Q1022**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1022-01	BP-VPB-190A-TB-20250103	1,2-Dichloroethane-d4	50	50.5	101	81	118
		Dibromofluoromethane	50	47.5	95	80	119
		Toluene-d8	50	50.9	102	89	112
		4-Bromofluorobenzene	50	51.8	104	85	114
Q1022-02	BP-VPB-190A-DUP-20250103	1,2-Dichloroethane-d4	50	51.3	103	81	118
		Dibromofluoromethane	50	47.1	94	80	119
		Toluene-d8	50	51.7	103	89	112
		4-Bromofluorobenzene	50	52.4	105	85	114
Q1022-03	BP-VPB-190A-GW-838-840	1,2-Dichloroethane-d4	50	50.3	101	81	118
		Dibromofluoromethane	50	47.5	95	80	119
		Toluene-d8	50	51.3	103	89	112
		4-Bromofluorobenzene	50	51.7	103	85	114
Q1022-04	BP-VPB-190A-GW-858-860	1,2-Dichloroethane-d4	50	50.5	101	81	118
		Dibromofluoromethane	50	45.8	92	80	119
		Toluene-d8	50	50.9	102	89	112
		4-Bromofluorobenzene	50	49.5	99	85	114
VX0107WBL01	VX0107WBL01	1,2-Dichloroethane-d4	50	50.6	101	81	118
		Dibromofluoromethane	50	47.1	94	80	119
		Toluene-d8	50	49.5	99	89	112
		4-Bromofluorobenzene	50	47.6	95	85	114
VX0107WBS02	VX0107WBS02	1,2-Dichloroethane-d4	50	50.2	100	81	118
		Dibromofluoromethane	50	48.5	97	80	119
		Toluene-d8	50	49.3	99	89	112
		4-Bromofluorobenzene	50	50.5	101	85	114
VX0107WBSD0	VX0107WBSD02	1,2-Dichloroethane-d4	50	49.7	99	81	118
		Dibromofluoromethane	50	46.3	93	80	119
		Toluene-d8	50	47.5	95	89	112
		4-Bromofluorobenzene	50	48.4	97	85	114