

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1022 SAS No.: Q1022 SDG NO.: Q1022
Lab File ID: VX044599.D Date Analyzed: 01/07/2025
Instrument ID: MSVOA_X Time Analyzed: 11:14
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	173796	5.54	298796	6.76	260263	10.05
UPPER LIMIT	347592	6.044	597592	7.257	520526	10.549
LOWER LIMIT	86898	5.044	149398	6.257	130132	9.549
EPA SAMPLE NO.						
BP-VPB-190A-TB-20250103	176098	5.54	318982	6.76	290099	10.05
BP-VPB-190A-DUP-20250103	172596	5.54	314042	6.76	292342	10.05
BP-VPB-190A-GW-838-840	188583	5.54	335673	6.76	308993	10.05
BP-VPB-190A-GW-858-860	178481	5.54	326468	6.76	293289	10.05
VX0107WBL01	173540	5.54	316021	6.75	270881	10.05
VX0107WBS02	220733	5.54	377450	6.75	324333	10.05
VX0107WBSD02	171730	5.54	300759	6.76	266797	10.05

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.