

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1031 SAS No.: Q1031 SDG No.: Q1031

Instrument ID: MSVOA_L Calibration Date/Time: 01/07/2025 09:18

Lab File ID: VL041838.D Init. Calib. Date(s): 01/07/2025 01/07/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:18 12:25

GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	RRF	RRF010	MIN RRF	%D	MAX%D
Vinyl Chloride	0.681	0.561		-17.62	
1,1-Dichloroethene	0.520	0.474		-8.85	
cis-1,2-Dichloroethene	1.103	1.078		-2.27	
1,1,1-Trichloroethane	1.804	1.685		-6.6	
Trichloroethene	0.367	0.351		-4.36	
Tetrachloroethene	0.283	0.290		2.47	
1-Bromo-4-Fluorobenzene	0.856	0.898		4.91	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.