

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010725\
 Data File : VL041843.D
 Acq On : 07 Jan 2025 11:52
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 01/09/2025
 Supervised By :Mahesh Dadoda 01/09/2025

Quant Time: Jan 08 01:23:18 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL010725AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Wed Jan 08 01:18:22 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.803	49	85525	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.984	114	192559	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.904	117	161678	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.396	95	131934	9.528	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.300%
Target Compounds						
5) Vinyl Chloride	1.586	62	224	0.038	ppbv #	32
32) 1,1,1-Trichloroethane	3.389	97	547	0.035	ppbv #	83
35) Carbon Tetrachloride	3.790	117	461m	0.032	ppbv	
63) 1,1,2,2-Tetrachloroethane	9.982	83	601m	0.036	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL010725\
Data File : VL041843.D
Acq On : 07 Jan 2025 11:52
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Jan 08 01:23:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL010725AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Jan 08 01:18:22 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 01/09/2025
Supervised By : Mahesh Dadoda 01/09/2025

