

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB134196

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3161
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/08/2025	14:25
2	CAL2	1	Water	NA	NA	20.2	7.00	01/08/2025	14:26
3	CAL3	1	Water	NA	NA	20.3	10.02	01/08/2025	14:30
4	ICV	1	Water	NA	NA	20.3	7.01	01/08/2025	14:33
5	CCV1	1	Water	NA	NA	20.3	2.01	01/08/2025	14:37
6	Q1032-01	1	Water	NA	NA	20.1	6.02	01/08/2025	14:45
7	Q1032-02	1	Water	NA	NA	20.3	7.06	01/08/2025	14:55
8	Q1032-03	1	Water	NA	NA	20.1	7.12	01/08/2025	15:00
9	Q1032-05	1	Water	NA	NA	20.6	5.73	01/08/2025	15:05
10	Q1032-06	1	Water	NA	NA	20.8	7.36	01/08/2025	15:15
11	Q1033-01	1	Water	NA	NA	20.1	7.39	01/08/2025	15:25
12	Q1036-01	1	Water	NA	NA	20.4	7.00	01/08/2025	15:30
13	Q1036-01DUP	1	Water	NA	NA	20.5	7.01	01/08/2025	15:31
14	CCV2	1	Water	NA	NA	20.2	12.02	01/08/2025	15:33

WORKLIST(Hardcopy Internal Chain)

VB 134196

WorkList Name : PH WATER P1036 WorkList ID : 186816 Department : Wet-Chemistry Date : 01-08-2025 14:06:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1032-01	C 50492	Water	pH	Cool 4 deg C	PSEG03	N41	01/08/2025	9040C
Q1032-02	0103-A	Water	pH	Cool 4 deg C	PSEG03	N41	01/08/2025	9040C
Q1032-03	0103-B	Water	pH	Cool 4 deg C	PSEG03	N41	01/08/2025	9040C
Q1032-05	0107-A	Water	pH	Cool 4 deg C	PSEG03	N41	01/08/2025	9040C
Q1032-06	0107-B	Water	pH	Cool 4 deg C	PSEG03	N41	01/08/2025	9040C
Q1033-01	R 28612	Water	pH	Cool 4 deg C	PSEG03	N41	01/08/2025	9040C
Q1036-01	E NP-WS-002	Water	pH	Cool 4 deg C	ENTA05	N51	01/08/2025	9040C

Date/Time 01-08-25 14:15
 Raw Sample Received by: RB WDC
 Raw Sample Relinquished by: RB WDC

Date/Time 01-08-25 16:30
 Raw Sample Received by: RB WDC
 Raw Sample Relinquished by: RB WDC