

LB134281


7470A

INSTRUMENT ID: CY1

Linear

μ Abs.:

41197



A= 0.0000e+000

B= 2.4141e-004

C= -1.3931e-001

Rho= 0.9990659

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.014	0.014	634	0.000	634					0/017
0.2	0.200	0.240	0.040	1571	0.0 %	1571					20
2.5	2.500	2.379	-0.121	10432	0.0 %	10432					-5
5.0	5.000	4.951	-0.049	21086	0.0 %	21086					4
7.5	7.500	7.810	0.310	32929	0.0 %	32929					2
10.0	10.000	9.806	-0.194	41197	0.0 %	41197					

LB134281 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 14 Jan 2025 13:31:00

Sample ID	Extended ID	Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	634	-	0.0000	7470A	PPB	14 Jan 2025 13:38:25	S	Std
0.2 - 1	50-2	1571	-	0.2000	7470A	PPB	14 Jan 2025 13:40:41	S	Std
2.5 - 1	52-5	10432	-	2.5000	7470A	PPB	14 Jan 2025 13:42:58	S	Std
5.0 - 1	55	21086	-	5.0000	7470A	PPB	14 Jan 2025 13:45:15	S	Std
7.5 - 1	57-5	32929	-	7.5000	7470A	PPB	14 Jan 2025 13:52:45	S	Std
10.0 - 1	514	41197	-	10.0000	7470A	PPB	14 Jan 2025 13:57:48	S	Std
ICV55 - 1	ICV55	15842	3.6851	-	7470A	PPB	14 Jan 2025 14:00:41	U	SMPL
ICB55 - 1	ICB55	412	-0.0399	-	7470A	PPB	14 Jan 2025 14:02:55	U	SMPL
CCV32 - 1	CCV32	21245	4.9895	-	7470A	PPB	14 Jan 2025 14:07:13	U	SMPL
CCB32 - 1	CCB32	494	-0.0201	-	7470A	PPB	14 Jan 2025 14:09:28	U	SMPL
CRA - 1	CRA	1448	0.2103	-	7470A	PPB	14 Jan 2025 14:14:40	U	SMPL
HighStd - 1	HighStd	40086	9.5379	-	7470A	PPB	14 Jan 2025 14:16:55	U	SMPL
ChkStd - 1	ChkStd	29320	6.9388	-	7470A	PPB	14 Jan 2025 14:19:10	U	SMPL
PB166050BL - 1	PBW	658	0.0195	-	7470A	PPB	14 Jan 2025 14:21:28	U	SMPL
PB166050BS - 1	LCSW	16009	3.7254	-	7470A	PPB	14 Jan 2025 14:23:47	U	SMPL
Q1036-02 - 1	NP-WS-002	912	0.0809	-	7470A	PPB	14 Jan 2025 14:26:03	U	SMPL
Q1036-02DUP - 1	NP-WS-002DUP	848	0.0654	-	7470A	PPB	14 Jan 2025 14:28:22	U	SMPL
Q1036-02MS - 1	NP-WS-002MS	16513	3.8471	-	7470A	PPB	14 Jan 2025 14:30:39	U	SMPL
Q1036-02MSD - 1	NP-WS-002MSD	15749	3.6627	-	7470A	PPB	14 Jan 2025 14:32:56	U	SMPL
PB166051BL - 1	PBW	857	0.0676	-	7470A	PPB	14 Jan 2025 14:35:13	U	SMPL
CCV33 - 1	CCV33	20915	4.9098	-	7470A	PPB	14 Jan 2025 14:37:30	U	SMPL
CCB33 - 1	CCB33	371	-0.0497	-	7470A	PPB	14 Jan 2025 14:39:46	U	SMPL
PB166051BS - 1	LCSW	15676	3.6450	-	7470A	PPB	14 Jan 2025 14:42:03	U	SMPL
Q1065-01 - 1	MOO-25-0002	-234	-0.1958	-	7470A	PPB	14 Jan 2025 14:44:18	U	SMPL
Q1069-02 - 1	RW7B-CARBON-20250109	965	0.0936	-	7470A	PPB	14 Jan 2025 14:46:35	U	SMPL
Q1074-02 - 1	HACKENSACK-DGA	523	-0.0131	-	7470A	PPB	14 Jan 2025 14:48:51	U	SMPL
Q1076-02 - 1	ARS20-0006	745	0.0405	-	7470A	PPB	14 Jan 2025 14:58:25	U	SMPL
Q1076-02DUP - 1	ARS20-0006DUP	778	0.0485	-	7470A	PPB	14 Jan 2025 15:00:41	U	SMPL
Q1076-02MS - 1	ARS20-0006MS	17256	4.0265	-	7470A	PPB	14 Jan 2025 15:02:57	U	SMPL
Q1076-02MSD - 1	ARS20-0006MSD	17222	4.0183	-	7470A	PPB	14 Jan 2025 15:05:14	U	SMPL
PB166025TB - 1	PB166025TB	737	0.0386	-	7470A	PPB	14 Jan 2025 15:07:33	U	SMPL
PB166026TB - 1	PB166026TB	931	0.0854	-	7470A	PPB	14 Jan 2025 15:09:50	U	SMPL
CCV34 - 1	CCV34	20017	4.6930	-	7470A	PPB	14 Jan 2025 15:12:05	U	SMPL
CCB34 - 1	CCB34	441	-0.0328	-	7470A	PPB	14 Jan 2025 15:14:21	U	SMPL
Q1036-02LX5 - 1		322	-0.0616	-	7470A	PPB	14 Jan 2025 15:16:39	U	SMPL
Q1036-02A - 1		14699	3.4092	-	7470A	PPB	14 Jan 2025 15:18:55	U	SMPL
Q1076-02LX5 - 1		69	-0.1227	-	7470A	PPB	14 Jan 2025 15:21:10	U	SMPL
Q1076-02A - 1		16853	3.9292	-	7470A	PPB	14 Jan 2025 15:23:27	U	SMPL
CCV35 - 1	CCV35	19447	4.5554	-	7470A	PPB	14 Jan 2025 15:25:43	U	SMPL
CCB35 - 1	CCB35	340	-0.0572	-	7470A	PPB	14 Jan 2025 15:28:02	U	SMPL
Q1076-02DUP - 1	ARS20-0006DUP	1019	0.1067	-	7470A	PPB	14 Jan 2025 14:53:23	U	SMPL