

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** ENTACT

**SDG No.:** Q1036

**Contract:** ENTA05

**Lab Code:** CHEM

**Case No.:** Q1036

**SAS No.:** Q1036

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte  | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |            |            |           |           |
|----------|-------------------------|------------------------------------------|------------|------------|-----------|-----------|
|          |                         | Al                                       | Ca         | Fe         | Mg        | Ag        |
| Aluminum | 396.100                 | 0.0000000                                | -0.0002060 | 0.0000000  | 0.0000000 | 0.0000000 |
| Arsenic  | 193.759                 | 0.0000000                                | 0.0000000  | -0.0000440 | 0.0000000 | 0.0000000 |
| Barium   | 493.409                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Cadmium  | 226.502                 | 0.0000000                                | 0.0000000  | 0.0000930  | 0.0000000 | 0.0000000 |
| Chromium | 267.716                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Cobalt   | 228.616                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Copper   | 224.700                 | 0.0000000                                | 0.0000000  | 0.0007850  | 0.0000000 | 0.0000000 |
| Iron     | 240.488                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Lead     | 220.353                 | -0.0000920                               | 0.0000000  | 0.0000380  | 0.0000000 | 0.0000000 |
| Nickel   | 231.604                 | 0.0000000                                | 0.0000000  | 0.0000000  | 0.0000000 | 0.0000000 |
| Selenium | 196.090                 | 0.0000000                                | 0.0000000  | -0.0001440 | 0.0000000 | 0.0000000 |
| Silver   | 328.068                 | 0.0000000                                | 0.0000000  | -0.0001490 | 0.0000000 | 0.0000000 |
| Zinc     | 213.800                 | 0.0000000                                | 0.0000000  | 0.0001050  | 0.0000000 | 0.0000000 |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** ENTACT

**SDG No.:** Q1036

**Contract:** ENTA05

**Lab Code:** CHEM

**Case No.:** Q1036

**SAS No.:** Q1036

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte  | Wave-<br>Length<br>(nm) | ICP Interement Correction Factors For: |           |           |           |            |
|----------|-------------------------|----------------------------------------|-----------|-----------|-----------|------------|
|          |                         | As                                     | Ba        | Be        | Cd        | Co         |
| Aluminum | 396.100                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Arsenic  | 193.759                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Barium   | 493.409                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Cadmium  | 226.502                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0002870  |
| Chromium | 267.716                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Cobalt   | 228.616                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Copper   | 224.700                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0009530  |
| Iron     | 240.488                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | -0.0039600 |
| Lead     | 220.353                 | 0.0000000                              | 0.0003170 | 0.0000000 | 0.0000000 | 0.0000000  |
| Nickel   | 231.604                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Selenium | 196.090                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | -0.0003570 |
| Silver   | 328.068                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Zinc     | 213.800                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** ENTACT

**SDG No.:** Q1036

**Contract:** ENTA05

**Lab Code:** CHEM

**Case No.:** Q1036

**SAS No.:** Q1036

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte  | Wave-<br>Length<br>(nm) | ICP Interement Correction Factors For: |           |           |           |            |
|----------|-------------------------|----------------------------------------|-----------|-----------|-----------|------------|
|          |                         | Cr                                     | Cu        | K         | Mn        | Mo         |
| Aluminum | 396.100                 | 0.0000000                              | 0.0000000 | 0.0000590 | 0.0000000 | 0.0396900  |
| Arsenic  | 193.759                 | -0.0029000                             | 0.0000000 | 0.0000000 | 0.0000000 | 0.0004900  |
| Barium   | 493.409                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Cadmium  | 226.502                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Chromium | 267.716                 | 0.0000000                              | 0.0000000 | 0.0000070 | 0.0002200 | 0.0000000  |
| Cobalt   | 228.616                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | -0.0007860 |
| Copper   | 224.700                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0006510 | 0.0020500  |
| Iron     | 240.488                 | 0.0000000                              | 0.0000000 | 0.0000730 | 0.0000000 | -0.0015250 |
| Lead     | 220.353                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0001400 | -0.0008600 |
| Nickel   | 231.604                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | 0.0000000  |
| Selenium | 196.090                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0007460 | 0.0000000  |
| Silver   | 328.068                 | 0.0000000                              | 0.0000000 | 0.0000000 | 0.0000000 | -0.0000120 |
| Zinc     | 213.800                 | 0.0000000                              | 0.0009010 | 0.0000000 | 0.0000000 | 0.0000000  |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** ENTACT

**SDG No.:** Q1036

**Contract:** ENTA05

**Lab Code:** CHEM

**Case No.:** Q1036

**SAS No.:** Q1036

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte  | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |            |           |           |           |
|----------|-------------------------|------------------------------------------|------------|-----------|-----------|-----------|
|          |                         | Na                                       | Ni         | Pb        | Sb        | Se        |
| Aluminum | 396.100                 | 0.0000000                                | 0.0000000  | 0.0012800 | 0.0000000 | 0.0000000 |
| Arsenic  | 193.759                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Barium   | 493.409                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Cadmium  | 226.502                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Chromium | 267.716                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Cobalt   | 228.616                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Copper   | 224.700                 | 0.0000000                                | -0.0047000 | 0.0036100 | 0.0000000 | 0.0000000 |
| Iron     | 240.488                 | 0.0000000                                | -0.0017000 | 0.0000000 | 0.0000000 | 0.0000000 |
| Lead     | 220.353                 | 0.0000000                                | 0.0006580  | 0.0000000 | 0.0000000 | 0.0001290 |
| Nickel   | 231.604                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Selenium | 196.090                 | 0.0000000                                | 0.0000000  | 0.0003330 | 0.0000000 | 0.0000000 |
| Silver   | 328.068                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Zinc     | 213.800                 | 0.0000000                                | 0.0067600  | 0.0000000 | 0.0000000 | 0.0000000 |

**Metals**

- 11 -

**ICP INTERELEMENT CORRECTION FACTORS**

**Client:** ENTACT

**SDG No.:** Q1036

**Contract:** ENTA05

**Lab Code:** CHEM

**Case No.:** Q1036

**SAS No.:** Q1036

**Instrument ID:** \_\_\_\_\_

**Date:** \_\_\_\_\_

**Interelement Correction Factors (apparent ppb analyte/ppm interferent )**

| Analyte  | Wave-<br>Length<br>(nm) | ICP Interelement Correction Factors For: |            |           |           |           |
|----------|-------------------------|------------------------------------------|------------|-----------|-----------|-----------|
|          |                         | Sn                                       | Ti         | Tl        | V         | Zn        |
| Aluminum | 396.100                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Arsenic  | 193.759                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Barium   | 493.409                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Cadmium  | 226.502                 | 0.0000000                                | 0.0000630  | 0.0001280 | 0.0000000 | 0.0000000 |
| Chromium | 267.716                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0001110 | 0.0000000 |
| Cobalt   | 228.616                 | 0.0000000                                | 0.0018800  | 0.0000000 | 0.0000000 | 0.0000000 |
| Copper   | 224.700                 | 0.0000000                                | 0.0003840  | 0.0000000 | 0.0000000 | 0.0000000 |
| Iron     | 240.488                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Lead     | 220.353                 | 0.0000000                                | -0.0003610 | 0.0000000 | 0.0000000 | 0.0000000 |
| Nickel   | 231.604                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Selenium | 196.090                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |
| Silver   | 328.068                 | 0.0000000                                | -0.0007420 | 0.0000000 | 0.0000000 | 0.0000000 |
| Zinc     | 213.800                 | 0.0000000                                | 0.0000000  | 0.0000000 | 0.0000000 | 0.0000000 |