

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q1036
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1036 **SAS No.:** Q1036
Matrix: Water **Sample ID:** Q1036-02 **Client ID:** NP-WS-002DUP
Percent Solids for Sample: NA **Duplicate ID** Q1036-02DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	500	U	500	U			P
Arsenic	ug/L	20	100	U	100	U			P
Barium	ug/L	20	79.8	J	82.6	J	3		P
Cadmium	ug/L	20	30.0	U	30.0	U			P
Chromium	ug/L	20	6.61	J	17.6	J	91		P
Cobalt	ug/L	20	150	U	150	U			P
Copper	ug/L	20	100	U	100	U			P
Iron	ug/L	20	500	U	500	U			P
Lead	ug/L	20	60.0	U	60.0	U			P
Mercury	ug/L	20	2.00	U	2.00	U			CV
Nickel	ug/L	20	9.69	J	20.1	J	70		P
Selenium	ug/L	20	100	U	100	U			P
Silver	ug/L	20	50.0	U	50.0	U			P
Zinc	ug/L	20	218		214		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: ENTACT **Level:** LOW **SDG No.:** Q1036
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1036 **SAS No.:** Q1036
Matrix: Water **Sample ID:** Q1036-02MS **Client ID:** NP-WS-002MSD
Percent Solids for Sample: NA **Duplicate ID** Q1036-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	9080		9050		0		P
Arsenic	ug/L	20	3820		3720		3		P
Barium	ug/L	20	979		964		2		P
Cadmium	ug/L	20	903		879		3		P
Chromium	ug/L	20	1960		1950		1		P
Cobalt	ug/L	20	936		915		2		P
Copper	ug/L	20	1470		1440		2		P
Iron	ug/L	20	14500		14200		2		P
Lead	ug/L	20	4530		4420		2		P
Mercury	ug/L	20	38.5		36.6		5		CV
Nickel	ug/L	20	2340		2280		3		P
Selenium	ug/L	20	9000		8780		2		P
Silver	ug/L	20	353		350		1		P
Zinc	ug/L	20	1210		1200		1		P