

Contract:	<u>ENTA05</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1036</u>	SAS No.:	<u>Q1036</u>	SDG NO.:	<u>Q1036</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):		<u>01/06/2025</u>	<u>01/07/2025</u>		
		Calibration Times:		<u>19:57</u>	<u>03:16</u>		

LAB FILE ID:		RT 1000 =	<u>PP068915.D</u>	RT 750 =	<u>PP068916.D</u>
RT 500 =	PP068917.D	RT 250 =	PP068918.D	RT 050 =	PP068919.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	10.53	10.52	10.53	10.53	10.53	10.53	10.43	10.63
Tetrachloro-m-xylene	4.67	4.67	4.68	4.67	4.67	4.67	4.57	4.77

Contract:	ENTA05					
Lab Code:	CHEM	Case No.:	Q1036	SAS No.:	Q1036	SDG NO.: Q1036
Instrument ID:	ECD_P	Calibration Date(s):	01/06/2025	01/07/2025		
		Calibration Times:	19:57	03:16		

LAB FILE ID:		RT 1000 =	<u>PP068915.D</u>	RT 750 =	<u>PP068916.D</u>
RT 500 =	PP068917.D	RT 250 =	PP068918.D	RT 050 =	PP068919.D

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Decachlorobiphenyl	9.11	9.11	9.11	9.11	9.11	9.11	9.01	9.21
Tetrachloro-m-xylene	3.98	3.99	3.98	3.99	3.99	3.99	3.89	4.09