

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1036 SAS No.: Q1036 SDG NO.: Q1036

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/06/2025 01/06/2025

Client Sample No.: CCAL02 Date Analyzed: 01/09/2025

Lab Sample No.: AR1660CCC500 Data File : PP069042.D Time Analyzed: 16:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.831	5.732	5.932	474.140	500.000	-5.2
Aroclor-1016-2	5.853	5.754	5.954	480.960	500.000	-3.8
Aroclor-1016-3	5.917	5.818	6.018	463.630	500.000	-7.3
Aroclor-1016-4	6.015	5.916	6.116	462.630	500.000	-7.5
Aroclor-1016-5	6.309	6.210	6.410	466.720	500.000	-6.7
Aroclor-1260-1	7.432	7.333	7.533	457.950	500.000	-8.4
Aroclor-1260-2	7.685	7.587	7.787	473.710	500.000	-5.3
Aroclor-1260-3	8.045	7.947	8.147	458.450	500.000	-8.3
Aroclor-1260-4	8.278	8.180	8.380	449.000	500.000	-10.2
Aroclor-1260-5	8.609	8.513	8.713	463.570	500.000	-7.3
Decachlorobiphenyl	10.517	10.424	10.624	44.640	50.000	-10.7
Tetrachloro-m-xylene	4.671	4.572	4.772	49.240	50.000	-1.5

CALIBRATION VERIFICATION SUMMARY

Contract: ENTA05

Lab Code: CHEM Case No.: Q1036 SAS No.: Q1036 SDG NO.: Q1036

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/06/2025 01/06/2025

Client Sample No.: CCAL02 Date Analyzed: 01/09/2025

Lab Sample No.: AR1660CCC500 Data File : PP069042.D Time Analyzed: 16:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.081	4.985	5.185	495.440	500.000	-0.9
Aroclor-1016-2	5.101	5.004	5.204	471.120	500.000	-5.8
Aroclor-1016-3	5.281	5.184	5.384	490.340	500.000	-1.9
Aroclor-1016-4	5.321	5.224	5.424	467.500	500.000	-6.5
Aroclor-1016-5	5.539	5.442	5.642	471.800	500.000	-5.6
Aroclor-1260-1	6.580	6.484	6.684	445.290	500.000	-10.9
Aroclor-1260-2	6.767	6.672	6.872	451.630	500.000	-9.7
Aroclor-1260-3	6.924	6.828	7.028	437.520	500.000	-12.5
Aroclor-1260-4	7.398	7.302	7.502	435.010	500.000	-13.0
Aroclor-1260-5	7.637	7.541	7.741	449.030	500.000	-10.2
Decachlorobiphenyl	9.106	9.011	9.211	41.870	50.000	-16.3
Tetrachloro-m-xylene	3.981	3.883	4.083	48.210	50.000	-3.6