

Surrogate Summary

SDG No.: Q1036

Client: ENTACT

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1036-01	NP-WS-002	1,2-Dichloroethane-d4	50	49.5	99	70 (74)	130 (125)
		Dibromofluoromethane	50	45.5	91	70 (75)	130 (124)
		Toluene-d8	50	50.3	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.5	101	70 (77)	130 (121)

Surrogate Summary

SDG No.: Q1036

Client: ENTACT

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VX0108WBL01	VX0108WBL01	1,2-Dichloroethane-d4	50	51.4	103	70 (74)	130 (125)
		Dibromofluoromethane	50	48.1	96	70 (75)	130 (124)
		Toluene-d8	50	50.9	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.7	109	70 (77)	130 (121)
VX0108WBS01	VX0108WBS01	1,2-Dichloroethane-d4	50	50.9	102	70 (74)	130 (125)
		Dibromofluoromethane	50	46.4	93	70 (75)	130 (124)
		Toluene-d8	50	48.8	98	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.5	97	70 (77)	130 (121)
VX0108WBSD0	VX0108WBSD01	1,2-Dichloroethane-d4	50	53.0	106	70 (74)	130 (125)
		Dibromofluoromethane	50	49.5	99	70 (75)	130 (124)
		Toluene-d8	50	50.1	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.5	99	70 (77)	130 (121)

