

Cover Page

Order ID : Q1037

Project ID : 63 Richards Street Dover, NJ

Client : All American Environmental, LLC

Lab Sample Number

Q1037-01
Q1037-02
Q1037-03

Client Sample Number

INDOOR
GARAGE
AMBIENT AIR

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 1/13/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

All American Environmental, LLC

Project Name: 63 Richards Street Dover, NJ

Project # N/A

Chemtech Project # Q1037

Test Name: TO-15

A. Number of Samples and Date of Receipt:

3 Air samples were received on 01/08/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested:
TO-15. This data package contains results for TO-15.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_L were done using GC column RTX-1, which is 60 meters, 0.32 mm id, 1.0 um df, Restek Cat. #10157. The Trap was supplied by Entech, glass bead and Tenax , Entech 7100A Preconcentrator. The analysis of TO-15 was based on method TO-15.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD for {Q1037-03DUP} with File ID: VL041875.D met criteria except for 2-Butanone[200%] due to difference in results of Original and DUP.

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

The Samples #INDOOR,GARAGE, AMBIENT AIR have the concentration of some target compounds below Method detection limits, therefore it is not reported as Hit in Form1.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

The Manual Integrations are performed for the followings.

Manual Integration Report			
Sequence	VL010925	Instrument	MSVOA_I

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
-----------	---------	-----------	-----------	-----------	---------------	---------------	--------

VSTDICCC010	VL041859.D	Ethanol			MMDADOD A	1/10/2025 7:22:34 AM	SAM	1/10/2025 7:26:21 AM	Peak Integrated by Software incorrectly
VSTDICCC010	VL041859.D	m/p-Xylene			MMDADOD A	1/10/2025 7:22:34 AM	SAM	1/10/2025 7:26:21 AM	Peak Integrated by Software incorrectly
VSTDICCC002	VL041860.D	1,1,2-Trichloroethane			MMDADOD A	1/10/2025 7:22:32 AM	SAM	1/10/2025 7:26:33 AM	Peak Integrated by Software incorrectly
VSTDICCC002	VL041860.D	1,2-Dichloropropane			MMDADOD A	1/10/2025 7:22:32 AM	SAM	1/10/2025 7:26:33 AM	Peak Integrated by Software incorrectly
VSTDICCC002	VL041860.D	Ethanol			MMDADOD A	1/10/2025 7:22:32 AM	SAM	1/10/2025 7:26:33 AM	Peak Integrated by Software incorrectly
VSTDICCC002	VL041860.D	m/p-Xylene			MMDADOD A	1/10/2025 7:22:32 AM	SAM	1/10/2025 7:26:33 AM	Peak Integrated by Software incorrectly
VSTDICCC002	VL041860.D	Methyl Methacrylate			MMDADOD A	1/10/2025 7:22:32 AM	SAM	1/10/2025 7:26:33 AM	Peak Integrated by Software incorrectly
VSTDICCC001	VL041861.D	1,4-Dioxane			MMDADOD A	1/10/2025 7:22:30 AM	SAM	1/10/2025 7:26:36 AM	Peak Integrated by

							Software incorrectl y
VSTDIC001	VL041861. D	2-Hexanone	MMDADOD A	1/10/202 5 7:22:30 AM	SA M	1/10/202 5 7:26:36 AM	Peak Integrate d by Software incorrectl y
VSTDIC001	VL041861. D	4-Methyl-2- Pentanone	MMDADOD A	1/10/202 5 7:22:30 AM	SA M	1/10/202 5 7:26:36 AM	Peak Integrate d by Software incorrectl y
VSTDIC001	VL041861. D	cis-1,3- Dichloropropene	MMDADOD A	1/10/202 5 7:22:30 AM	SA M	1/10/202 5 7:26:36 AM	Peak Integrate d by Software incorrectl y
VSTDIC001	VL041861. D	Dibromochloromethan e	MMDADOD A	1/10/202 5 7:22:30 AM	SA M	1/10/202 5 7:26:36 AM	Peak Integrate d by Software incorrectl y
VSTDIC001	VL041861. D	Ethanol	MMDADOD A	1/10/202 5 7:22:30 AM	SA M	1/10/202 5 7:26:36 AM	Peak Integrate d by Software incorrectl y
VSTDIC001	VL041861. D	m/p-Xylene	MMDADOD A	1/10/202 5 7:22:30 AM	SA M	1/10/202 5 7:26:36 AM	Peak Integrate d by Software incorrectl y
VSTDIC001	VL041861. D	t-1,3-Dichloropropene	MMDADOD A	1/10/202 5 7:22:30 AM	SA M	1/10/202 5 7:26:36 AM	Peak Integrate d by Software incorrectl y
VSTDIC001	VL041861. D	trans-1,2- Dichloroethene	MMDADOD A	1/10/202 5 7:22:30 AM	SA M	1/10/202 5 7:26:36 AM	Peak Integrate d by Software incorrectl y
VSTDIC0.5	VL041862. D	1,1,2-Trichloroethane	MMDADOD A	1/10/202 5 7:22:28 AM	SA M	1/10/202 5 7:26:38 AM	Peak Integrate d by Software incorrectl

							y
VSTDIC0.5	VL041862.D	1,2-Dichloropropane	MMDADOD A	1/10/2025 7:22:28 AM	SAM	1/10/2025 7:26:38 AM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL041862.D	1,4-Dioxane	MMDADOD A	1/10/2025 7:22:28 AM	SAM	1/10/2025 7:26:38 AM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL041862.D	2,2,4-Trimethylpentane	MMDADOD A	1/10/2025 7:22:28 AM	SAM	1/10/2025 7:26:38 AM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL041862.D	4-Methyl-2-Pentanone	MMDADOD A	1/10/2025 7:22:28 AM	SAM	1/10/2025 7:26:38 AM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL041862.D	Cyclohexane	MMDADOD A	1/10/2025 7:22:28 AM	SAM	1/10/2025 7:26:38 AM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL041862.D	Ethanol	MMDADOD A	1/10/2025 7:22:28 AM	SAM	1/10/2025 7:26:38 AM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL041862.D	Heptane	MMDADOD A	1/10/2025 7:22:28 AM	SAM	1/10/2025 7:26:38 AM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL041862.D	m/p-Xylene	MMDADOD A	1/10/2025 7:22:28 AM	SAM	1/10/2025 7:26:38 AM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL041862.D	Methyl Methacrylate	MMDADOD A	1/10/2025 7:22:28 AM	SAM	1/10/2025 7:26:38 AM	Peak Integrated by Software incorrectly
VSTDIC0.5	VL041862.D	Trichloroethene	MMDADOD	1/10/202	SA	1/10/202	Peak

	D		A	5 7:22:28 AM	M	5 7:26:38 AM	Integrate d by Software incorrectl y
VSTDIC0.5	VL041862. D	Vinyl Acetate	MMDADOD A	1/10/202 5 7:22:28 AM	SA M	1/10/202 5 7:26:38 AM	Peak Integrate d by Software incorrectl y
VSTDIC0.1	VL041863. D	1,1,1-Trichloroethane	MMDADOD A	1/10/202 5 7:22:26 AM	SA M	1/10/202 5 7:26:40 AM	Peak Integrate d by Software incorrectl y
VSTDIC0.1	VL041863. D	1,2-Dibromoethane	MMDADOD A	1/10/202 5 7:22:26 AM	SA M	1/10/202 5 7:26:40 AM	Peak Integrate d by Software incorrectl y
VSTDIC0.1	VL041863. D	Carbon Tetrachloride	MMDADOD A	1/10/202 5 7:22:26 AM	SA M	1/10/202 5 7:26:40 AM	Peak Integrate d by Software incorrectl y
VSTDIC0.1	VL041863. D	Tetrachloroethene	MMDADOD A	1/10/202 5 7:22:26 AM	SA M	1/10/202 5 7:26:40 AM	Peak Integrate d by Software incorrectl y
VSTDIC0.1	VL041863. D	Trichloroethene	MMDADOD A	1/10/202 5 7:22:26 AM	SA M	1/10/202 5 7:26:40 AM	Peak Integrate d by Software incorrectl y
VSTDIC0.0 3	VL041864. D	1,1,1-Trichloroethane	MMDADOD A	1/10/202 5 7:22:24 AM	SA M	1/10/202 5 7:26:42 AM	Peak Integrate d by Software incorrectl y
VSTDIC0.0 3	VL041864. D	Carbon Tetrachloride	MMDADOD A	1/10/202 5 7:22:24 AM	SA M	1/10/202 5 7:26:42 AM	Peak Integrate d by Software incorrectl y
VSTDIC0.0 3	VL041864. D	Trichloroethene	MMDADOD A	1/10/202 5 7:22:24 AM	SA M	1/10/202 5 7:26:42 AM	Peak Integrate d by

							Software incorrectl y
VSTDIC015	VL041865. D	Ethanol	MMDADOD A	1/10/202 5 7:22:22 AM	SA M	1/10/202 5 7:26:44 AM	Peak Integrate d by Software incorrectl y
VSTDIC015	VL041865. D	m/p-Xylene	MMDADOD A	1/10/202 5 7:22:22 AM	SA M	1/10/202 5 7:26:44 AM	Peak Integrate d by Software incorrectl y
VSTDICV010	VL041866. D	Ethanol	MMDADOD A	1/10/202 5 7:22:20 AM	SA M	1/10/202 5 7:26:46 AM	Peak Integrate d by Software incorrectl y
VSTDICV010	VL041866. D	m/p-Xylene	MMDADOD A	1/10/202 5 7:22:20 AM	SA M	1/10/202 5 7:26:46 AM	Peak Integrate d by Software incorrectl y
VL0109ABS0 1	VL041870. D	m/p-Xylene	MMDADOD A	1/10/202 5 7:22:18 AM	SA M	1/10/202 5 7:26:48 AM	Peak Integrate d by Software incorrectl y
VL0109ABS0 1	VL041870. D	Trichloroethene	MMDADOD A	1/10/202 5 7:22:18 AM	SA M	1/10/202 5 7:26:48 AM	Peak Integrate d by Software incorrectl y
Q1037-01	VL041872. D	Carbon Tetrachloride	MMDADOD A	1/10/202 5 7:22:16 AM	SA M	1/10/202 5 7:26:51 AM	Peak Integrate d by Software incorrectl y
Q1037-01	VL041872. D	Cyclohexane	MMDADOD A	1/10/202 5 7:22:16 AM	SA M	1/10/202 5 7:26:51 AM	Peak Integrate d by Software incorrectl y
Q1037-01	VL041872. D	m/p-Xylene	MMDADOD A	1/10/202 5 7:22:16 AM	SA M	1/10/202 5 7:26:51 AM	Peak Integrate d by Software incorrectl

							y
Q1037-01	VL041872.D	Propene	MMDADOD A	1/10/2025 7:22:16 AM	SAM	1/10/2025 7:26:51 AM	Peak Integrated by Software incorrectly
Q1037-01	VL041872.D	Toluene	MMDADOD A	1/10/2025 7:22:16 AM	SAM	1/10/2025 7:26:51 AM	Peak Integrated by Software incorrectly
Q1037-02	VL041873.D	Benzene	MMDADOD A	1/10/2025 7:22:14 AM	SAM	1/10/2025 7:26:53 AM	Peak Integrated by Software incorrectly
Q1037-02	VL041873.D	Carbon Tetrachloride	MMDADOD A	1/10/2025 7:22:14 AM	SAM	1/10/2025 7:26:53 AM	Peak Integrated by Software incorrectly
Q1037-02	VL041873.D	Propene	MMDADOD A	1/10/2025 7:22:14 AM	SAM	1/10/2025 7:26:53 AM	Peak Integrated by Software incorrectly
Q1037-03	VL041874.D	Carbon Tetrachloride	MMDADOD A	1/10/2025 7:22:11 AM	SAM	1/10/2025 7:26:55 AM	Peak Integrated by Software incorrectly
Q1037-03	VL041874.D	Propene	MMDADOD A	1/10/2025 7:22:11 AM	SAM	1/10/2025 7:26:55 AM	Peak Integrated by Software incorrectly

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1037

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: MOHAMMAD AHMED

Date: 01/13/2025