

DATA PACKAGE

VOLATILE ORGANICS

PROJECT NAME : WASTE WATER 2025

GARDEN STATE LABORATORIES, INC.

410 Hillside Avenue

Hillside, NJ - 07205

Phone No: 800-273-8901

ORDER ID : Q1038

ATTENTION : Sharon Ercoliani



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) VOCMS Group1- Case Narrative	4
2.2) VOCMS Group2- Case Narrative	6
3) Qualifier Page	8
4) QA Checklist	9
5) VOCMS Group1 Data	10
6) VOCMS Group2 Data	15
7) Shipping Document	24
7.1) CHAIN OF CUSTODY	25
7.2) Lab Certificate	26
7.3) Internal COC	27

1
2
3
4
5
6
7

Cover Page

Order ID : Q1038

Project ID : Waste Water 2025

Client : Garden State Laboratories, Inc.

Lab Sample Number

Q1038-01
Q1038-02

Client Sample Number

250108071-01-VOA
250108055-07-TRIP-BLANK

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 10:41 am, Jan 21, 2025

Date: 1/16/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Garden State Laboratories, Inc.

Project Name: Waste Water 2025

Project # N/A

Chemtech Project # Q1038

Test Name: VOCMS Group1

A. Number of Samples and Date of Receipt:

2 Water samples were received on 01/09/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOCMS Group1 and VOCMS Group2. This data package contains results for VOCMS Group1.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOCMS Group1 was based on method 624.1.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

The pH value of the samples was 6.0 as samples received unpreserved.

As per method, MS/MSD is required to be performed with the sample analysis. However, Lab did not receive sufficient volume to perform the MS/MSD therefore MS/MSD were not performed for this project. However, Lab has performed LCS/LCSD instead.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <35% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount

for all compounds using Linear Regression when the %RSD value for a compound is > 35% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature _____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 10:42 am, Jan 21, 2025

CASE NARRATIVE

Garden State Laboratories, Inc.

Project Name: Waste Water 2025

Project # N/A

Chemtech Project # Q1038

Test Name: VOCMS Group2

A. Number of Samples and Date of Receipt:

2 Water samples were received on 01/09/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: VOCMS Group1 and VOCMS Group2. This data package contains results for VOCMS Group2.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_N were done using GC column Rxi-624SIL MS 30m, 0.25mm, 1.4 um, Cat. #13868. The analysis of VOCMS Group2 was based on method 8260D.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank Spike Duplicate met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

The pH value of the samples was 6.0 as samples received unpreserved.

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount

for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature _____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 10:42 am, Jan 21, 2025

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: <ul style="list-style-type: none"> (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1038

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: MOHAMMAD AHMED

Date: 01/16/2025

Hit Summary Sheet
624.1

SDG No.: Q1038

Client: Garden State Laboratories, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID:

0

Total Voc :

Total Concentration:

A

B

C

D



SAMPLE DATA

Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	01/08/25	
Project:	Waste Water 2025		Date Received:	01/09/25	
Client Sample ID:	250108071-01-VOA		SDG No.:	Q1038	
Lab Sample ID:	Q1038-01		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085423.D	1		01/10/25 13:20	VN011025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
107-02-8	Acrolein	9.30	U	9.30	25.0	ug/L
107-13-1	Acrylonitrile	3.70	U	3.70	25.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.1		91 - 110	100%	SPK: 30
2037-26-5	Toluene-d8	27.3		91 - 112	91%	SPK: 30
460-00-4	4-Bromofluorobenzene	24.8		63 - 112	83%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	27800	7.812			
540-36-3	1,4-Difluorobenzene	144000	9.094			
3114-55-4	Chlorobenzene-d5	130000	11.865			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	01/08/25	
Project:	Waste Water 2025		Date Received:	01/09/25	
Client Sample ID:	250108055-07-TRIP-BLANK		SDG No.:	Q1038	
Lab Sample ID:	Q1038-02		Matrix:	Water	
Analytical Method:	E624.1		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085422.D	1		01/10/25 12:56	VN011025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
107-02-8	Acrolein	9.30	U	9.30	25.0	ug/L
107-13-1	Acrylonitrile	3.70	U	3.70	25.0	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	30.0		91 - 110	100%	SPK: 30
2037-26-5	Toluene-d8	28.6		91 - 112	95%	SPK: 30
460-00-4	4-Bromofluorobenzene	24.4		63 - 112	81%	SPK: 30
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	38700	7.806			
540-36-3	1,4-Difluorobenzene	199000	9.1			
3114-55-4	Chlorobenzene-d5	178000	11.865			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q1038	OrderDate:	1/9/2025 9:48:00 AM
Client:	Garden State Laboratories, Inc.	Project:	Waste Water 2025
Contact:	Sharon Ercoliani	Location:	VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1038-01	250108071-01-VOA	Water	VOCMS Group1	624.1	01/08/25		01/10/25	01/09/25
Q1038-02	250108055-07-TRIP-BLANK	Water	VOCMS Group1	624.1	01/08/25		01/10/25	01/09/25

Hit Summary Sheet
SW-846

SDG No.: Q1038
Client: Garden State Laboratories, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID: 250108071-01-VOA								
Q1038-01	250108071-01-VOA	Water	Methyl tert-butyl Ether	0.36	J	0.16	1.00	ug/L
Q1038-01	250108071-01-VOA	Water	Methylene Chloride	0.44	J	0.32	1.00	ug/L
Q1038-01	250108071-01-VOA	Water	Benzene	0.44	J	0.16	1.00	ug/L
Q1038-01	250108071-01-VOA	Water	Chlorobenzene	4.00		0.13	1.00	ug/L
Q1038-01	250108071-01-VOA	Water	o-Xylene	0.34	J	0.14	1.00	ug/L
Q1038-01	250108071-01-VOA	Water	1,4-Dichlorobenzene	4.80		0.27	1.00	ug/L
Q1038-01	250108071-01-VOA	Water	1,2-Dichlorobenzene	0.36	J	0.19	1.00	ug/L
			Total Voc :			10.7		
Q1038-01	250108071-01-VOA	Water	Tetrahydrofuran	* 110	J	1.20	5.00	ug/L
Q1038-01	250108071-01-VOA	Water	Tert butyl alcohol	* 180	J	5.60	25.0	ug/L
Q1038-01	250108071-01-VOA	Water	Diethyl Ether	* 8.00	J	0.20	1.00	ug/L
Q1038-01	250108071-01-VOA	Water	Naphthalene	* 0.96	J	0.59	1.00	ug/L
Q1038-01	250108071-01-VOA	Water	1,4-Dioxane	* 140	J	6.50	100	ug/L
			Total Tics :			439		
			Total Concentration:			450		
Client ID: 250108055-07-TRIP-BLANK								
Q1038-02	250108055-07-TRII	Water	Naphthalene	* 2.80	J	0.59	1.00	ug/L
			Total Tics :			2.80		
			Total Concentration:			2.80		



SAMPLE DATA

Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	01/08/25	
Project:	Waste Water 2025		Date Received:	01/09/25	
Client Sample ID:	250108071-01-VOA		SDG No.:	Q1038	
Lab Sample ID:	Q1038-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group2	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085409.D	1		01/09/25 15:15	VN010925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	1.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	1.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	1.00	ug/L
67-64-1	Acetone	1.40	U	1.40	5.00	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.36	J	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	1.00	ug/L
75-09-2	Methylene Chloride	0.44	J	0.32	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	1.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	1.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	1.00	ug/L
71-43-2	Benzene	0.44	J	0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	1.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	1.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	5.00	ug/L
108-88-3	Toluene	0.18	U	0.18	1.00	ug/L

Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	01/08/25	
Project:	Waste Water 2025		Date Received:	01/09/25	
Client Sample ID:	250108071-01-VOA		SDG No.:	Q1038	
Lab Sample ID:	Q1038-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group2	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085409.D	1		01/09/25 15:15	VN010925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	1.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	1.00	ug/L
108-90-7	Chlorobenzene	4.00		0.13	1.00	ug/L
100-41-4	Ethyl Benzene	0.16	U	0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	2.00	ug/L
95-47-6	o-Xylene	0.34	J	0.14	1.00	ug/L
100-42-5	Styrene	0.16	U	0.16	1.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	1.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	4.80		0.27	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.36	J	0.19	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.4		74 - 125	107%	SPK: 50
1868-53-7	Dibromofluoromethane	53.3		75 - 124	107%	SPK: 50
2037-26-5	Toluene-d8	52.1		86 - 113	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.1		77 - 121	98%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	181000	8.224			
540-36-3	1,4-Difluorobenzene	323000	9.1			
3114-55-4	Chlorobenzene-d5	281000	11.859			
3855-82-1	1,4-Dichlorobenzene-d4	116000	13.788			
TENTATIVE IDENTIFIED COMPOUNDS						

Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	01/08/25	
Project:	Waste Water 2025		Date Received:	01/09/25	
Client Sample ID:	250108071-01-VOA		SDG No.:	Q1038	
Lab Sample ID:	Q1038-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group2	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085409.D	1		01/09/25 15:15	VN010925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
60-29-7	Diethyl Ether	8.00	J		3.95	ug/L
75-65-0	Tert butyl alcohol	180	J		5.51	ug/L
109-99-9	Tetrahydrofuran	110	J		7.84	ug/L
123-91-1	1,4-Dioxane	140	J		9.69	ug/L
91-20-3	Naphthalene	0.96	J		15.6	ug/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Garden State Laboratories, Inc.	Date Collected:	01/08/25
Project:	Waste Water 2025	Date Received:	01/09/25
Client Sample ID:	250108055-07-TRIP-BLANK	SDG No.:	Q1038
Lab Sample ID:	Q1038-02	Matrix:	Water
Analytical Method:	SW8260	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group2
GC Column:	RXI-624 ID : 0.25	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085408.D	1		01/09/25 14:51	VN010925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.21	U	0.21	1.00	ug/L
74-87-3	Chloromethane	0.35	U	0.35	1.00	ug/L
75-01-4	Vinyl Chloride	0.34	U	0.34	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.56	U	0.56	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.34	U	0.34	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.26	U	0.26	1.00	ug/L
67-64-1	Acetone	1.40	U	1.40	5.00	ug/L
75-15-0	Carbon Disulfide	0.32	U	0.32	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.60	U	0.60	1.00	ug/L
75-09-2	Methylene Chloride	0.32	U	0.32	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.60	U	1.60	5.00	ug/L
78-93-3	2-Butanone	1.30	U	1.30	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.25	U	0.25	1.00	ug/L
74-97-5	Bromochloromethane	0.18	U	0.18	1.00	ug/L
67-66-3	Chloroform	0.26	U	0.26	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.19	U	0.19	1.00	ug/L
108-87-2	Methylcyclohexane	0.19	U	0.19	1.00	ug/L
71-43-2	Benzene	0.16	U	0.16	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.24	U	0.24	1.00	ug/L
79-01-6	Trichloroethene	0.32	U	0.32	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.19	U	0.19	1.00	ug/L
75-27-4	Bromodichloromethane	0.24	U	0.24	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.75	U	0.75	5.00	ug/L
108-88-3	Toluene	0.18	U	0.18	1.00	ug/L

Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	01/08/25	
Project:	Waste Water 2025		Date Received:	01/09/25	
Client Sample ID:	250108055-07-TRIP-BLANK		SDG No.:	Q1038	
Lab Sample ID:	Q1038-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group2	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085408.D	1		01/09/25 14:51	VN010925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.21	U	0.21	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.18	U	0.18	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	1.10	U	1.10	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.16	U	0.16	1.00	ug/L
127-18-4	Tetrachloroethene	0.25	U	0.25	1.00	ug/L
108-90-7	Chlorobenzene	0.13	U	0.13	1.00	ug/L
100-41-4	Ethyl Benzene	0.16	U	0.16	1.00	ug/L
179601-23-1	m/p-Xylenes	0.31	U	0.31	2.00	ug/L
95-47-6	o-Xylene	0.14	U	0.14	1.00	ug/L
100-42-5	Styrene	0.16	U	0.16	1.00	ug/L
75-25-2	Bromoform	0.21	U	0.21	1.00	ug/L
98-82-8	Isopropylbenzene	0.13	U	0.13	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.27	U	0.27	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.24	U	0.24	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.27	U	0.27	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.46	U	0.46	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.42	U	0.42	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.51	U	0.51	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	55.7		74 - 125	111%	SPK: 50
1868-53-7	Dibromofluoromethane	52.7		75 - 124	105%	SPK: 50
2037-26-5	Toluene-d8	52.0		86 - 113	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.4		77 - 121	95%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	172000	8.218			
540-36-3	1,4-Difluorobenzene	318000	9.1			
3114-55-4	Chlorobenzene-d5	277000	11.865			
3855-82-1	1,4-Dichlorobenzene-d4	110000	13.788			
TENTATIVE IDENTIFIED COMPOUNDS						

Report of Analysis

Client:	Garden State Laboratories, Inc.		Date Collected:	01/08/25	
Project:	Waste Water 2025		Date Received:	01/09/25	
Client Sample ID:	250108055-07-TRIP-BLANK		SDG No.:	Q1038	
Lab Sample ID:	Q1038-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group2	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085408.D	1		01/09/25 14:51	VN010925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
91-20-3	Naphthalene	2.80	J		15.6	ug/L

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

LAB CHRONICLE

OrderID:	Q1038	OrderDate:	1/9/2025 9:48:00 AM
Client:	Garden State Laboratories, Inc.	Project:	Waste Water 2025
Contact:	Sharon Ercoliani	Location:	VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1038-01	250108071-01-VOA	Water			01/08/25			01/09/25
			VOCMS Group1	624.1				
			VOCMS Group2	8260-Low				
Q1038-02	250108055-07-TRIP-BLANK	Water			01/08/25			01/09/25
			VOCMS Group1	624.1				
			VOCMS Group2	8260-Low				



SHIPPING DOCUMENTS

Garden State Laboratories, Inc.

Main Lab - 410 Hillside Avenue, Hillside NJ 07205 - NJDEP Lab Cert. #20044
Jersey Shore Lab - 54 Main Street, Waretown NJ 08758 - NJDEP Lab Cert. #15037

Tel. 800-273-8901/908-688-8900 Fax 908-688-8966 www.gslabs.com info@gslabs.com

Office and Drop off Locations

North Jersey Office: 225 Sparta Avenue, Sparta, NJ 07871 Tel. 973-729-1827

West Jersey Office: 2050 Route 31 North, Glen Gardner, NJ 08826 Tel. 908-537-7414

CLIENT INFORMATION (REPORT TO BE SENT TO)

Name: Garden State Laboratories, Inc. Contact/Authorized by: Elinor Battler
 Mailing Address: 410 Hillside Ave. Phone: 908-688-8900 x 303
 City/State/Zip: Hillside, NJ 07205 Email: ebattler@gslabs.com

SAMPLE INFORMATION

SAMPLE TYPE: WASTE WATER
 SAMPLE LOCATION: ACUA SW LANDFILL LEACHATE TANKS

Q1038

OR SAMPLE RECEIVING USE ONLY

DATE/TIME/TEMP. REC'D AT LAB:

Page of

GSL CLIENT #

MICRO #

CHEM. #

SAMPLE REC'D BY:

☒ GSL FIELD SAMPLER/PICK-UP

☐ PICK-UP AT DROP OFF LOCATION

☐ DELIVERED BY CLIENT

Grab/Comp	SAMPLE ID	SAMPLE COLLECTION				ANALYSIS REQUIRED (Print Legibly)		CONTAINER INFORMATION			
		Date	Time	AM	PM	<input type="checkbox"/> List attached	Total Pages	No.	Type*	Size	Pres.*
X	VOA 250108071-01	1/8/25	10:18	X		EPA 8260 TCL LIST + Acrolien & Acrylonitrile		3	V	40ml	A
X	Trip blank 250108055-01	—	—			EPA 8260 TCL LIST + Acrolien & Aci		2	V	40ml	A

*Container type: P = Plastic G = Glass A = Amber Glass T = Sterile Thio V = Vial Other/Specify:

*Preservation Code: A = Non Preserved B = Sulfuric Acid C = Sodium Hydroxide D = Nitric Acid
 E = Hydrochloric Acid F = Zinc Acetate G = Sodium Thiosulfate H = Ascorbic Acid I = Cooled Other/Specify:

☒ SUBCONTRACTED WORK

TURNAROUND TIME: ☒ Stand ☐ Rush (IF RUSH REQUESTED) Rush Due by:

REPORT FORMAT: ☒ Standard Report ☐ Other/Specify:

Standard Report + E2 PWS ID#:

SEND TO: Chem Tech

DATE/TIME: 1-9-25-0945

METHOD OF SHIPMEN Deliver

PAYMENT INFORMATION

☐ Sampling/Pick-up Fee: \$ ☐ Composite Fee: \$ ☐ Rush Fee: \$ Amount Due: \$

Payment Method: ☐ Credit Card Type: ☐ Check # ☐ Other: See Quote

Note:

VOA UNPRESERVED DUE TO EFFERVESCENCE - 3 DAY TAT PER JORDAN HE

SAMPLE CUSTODY EXCHANGES MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION
PLEASE PRINT YOUR NAME LEGIBLY, USE FULL LEGAL SIGNATURE, DATE AND TIME

Sampled by (PRINT):	Signature:	Date/Time:
Client/Client's Representative (PRINT):	Signature:	Date/Time:
1. Received/Relinquished by (PRINT): Daniel Askey	Signature: Daniel Askey	Date/Time: 1/8/25 15:22
2. Received/Relinquished by (PRINT): GUYTON WHITE	Signature: GUYTON WHITE	Date/Time: 1-9-25-0945

The liability of Garden State Laboratories, Inc. for services rendered shall in no event exceed the amount of the invoice.
 Main Lab certified by NJ Dept. of Health, NJDEP-TNI, NY Dept. of Health #11550 and PADEP #68-03680

CHAIN OF CUSTODY RECORD - PRESS HARD AND PRINT CLEARLY - USE BALL POINT PEN

IMPORTANT: PRINTED NAMES & SIGNATURES ARE REQUIRED


Laboratory Certification


Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1038	GARD04	Order Date : 1/9/2025 9:48:00 AM	Project Mgr :
Client Name : Garden State Laboratories, I		Project Name : Waste Water 2025	Report Type : Level 1
Client Contact : Sharon Ercoliani		Receive DateTime : 1/9/2025 9:47:00 AM	EDD Type : EXCEL NOCLEANUP
Invoice Name : Garden State Laboratories, I		Purchase Order :	Hard Copy Date :
Invoice Contact : Sharon Ercoliani			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1038-01	250108071-01-VOA	Water	01/08/2025	10:18					
					VOCMS Group1		624.1	10 Bus. Days	
					VOCMS Group2		8260-Low	10 Bus. Days	
Q1038-02	250108055-07-TRIP-BLANK	Water	01/08/2025	00:00					
					VOCMS Group1		624.1	10 Bus. Days	
					VOCMS Group2		8260-Low	10 Bus. Days	

Relinquished By : 
Date / Time : 1/9/25 1040

Received By : 
Date / Time : 1/9/25 10:40

Storage Area : VOA Refridgerator Room