

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GARD04

Lab Code: CHEM Case No.: Q1038 SAS No.: Q1038 SDG No.: Q1038

Instrument ID: MSVOA_N Calibration Date/Time: 01/10/2025 10:56

Lab File ID: VN085418.D Init. Calib. Date(s): 12/20/2024 12/20/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:05 11:42

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF020	MIN RRF	%D	MAX%D
Acrolein	0.247	0.249		0.81	
Acrylonitrile	1.070	0.923		-13.74	
1,2-Dichloroethane-d4	2.639	2.574	0.01	-2.46	
Toluene-d8	1.465	1.436	0.01	-1.98	
4-Bromofluorobenzene	0.486	0.454	0.2	-6.58	

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.