

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1040
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1040 **SAS No.:** Q1040
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	9620	10000	96	90 - 110	P	01/17/2025	11:21	LB134334

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1040
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1040 **SAS No.:** Q1040
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	102	100	102	80 - 120	P	01/17/2025	11:26	LB134334

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1040
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1040 **SAS No.:** Q1040
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5080	5000	102	90 - 110	P	01/17/2025	11:56	LB134334
CCV02	Iron	4910	5000	98	90 - 110	P	01/17/2025	12:47	LB134334
CCV03	Iron	4900	5000	98	90 - 110	P	01/17/2025	13:36	LB134334
CCV04	Iron	4940	5000	99	90 - 110	P	01/17/2025	14:27	LB134334
CCV05	Iron	4740	5000	95	90 - 110	P	01/17/2025	15:34	LB134334
CCV06	Iron	4940	5000	99	90 - 110	P	01/17/2025	16:16	LB134334
CCV07	Iron	4770	5000	96	90 - 110	P	01/17/2025	17:11	LB134334
CCV08	Iron	4960	5000	99	90 - 110	P	01/17/2025	17:52	LB134334