

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tetra Tech NUS, Inc. **SDG No.:** Q1040
Contract: TETR06 **Lab Code:** CHEM **Case No.:** Q1040 **SAS No.:** Q1040
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	95900	101000	95	85600	116500	01/17/2025	11:39	LB134334
ICSAB01	Iron	101000	99300	102	84400	114500	01/17/2025	11:43	LB134334