

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Tetra Tech NUS, Inc. **level:** low **sdg no.:** Q1040
contract: TETR06 **lab code:** CHEM **case no.:** Q1040 **sas no.:** Q1040
matrix: Water **sample id:** Q1048-01 **client id:** WATER TREATMENT DISCHARGEMS
Percent Solids for Sample: NA **Spiked ID:** Q1048-01MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	2760		1180		1500	105		P

metals

- 5a -

MATRIX SPIKE DUPLICATE SUMMARY

client: Tetra Tech NUS, Inc. **level:** low **sdg no.:** Q1040
contract: TETR06 **lab code:** CHEM **case no.:** Q1040 **sas no.:** Q1040
matrix: Water **sample id:** Q1048-01 **client id:** WATER TREATMENT DISCHARGEMSD
Percent Solids for Sample: NA **Spiked ID:** Q1048-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Iron	ug/L	87 - 115	2730		1180		1500	104		P