

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011025\
 Data File : BN035904.D
 Acq On : 10 Jan 2025 13:17
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 10 18:06:40 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	121	0.00
2	1,4-Dioxane	0.397	0.432	-8.8	140	0.00
3	n-Nitrosodimethylamine	0.693	0.746	-7.6	133	-0.01
4 S	2-Fluorophenol	0.981	0.965	1.6	123	0.00
5 S	Phenol-d6	1.219	1.165	4.4	119	0.00
6	bis(2-Chloroethyl)ether	0.929	0.946	-1.8	122	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
8 S	Nitrobenzene-d5	0.317	0.347	-9.5	129	-0.01
9	Naphthalene	1.123	1.118	0.4	113	0.00
10	Hexachlorobutadiene	0.365	0.370	-1.4	115	0.00
11 SURR	2-Methylnaphthalene-d10	0.536	0.503	6.2	105	0.00
12	2-Methylnaphthalene	0.695	0.661	4.9	106	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.192	0.204	-6.2	112	-0.01
15 S	2-Fluorobiphenyl	1.755	1.724	1.8	104	0.00
16	Acenaphthylene	1.884	1.763	6.4	100	0.00
17	Acenaphthene	1.235	1.199	2.9	103	0.00
18	Fluorene	1.359	1.276	6.1	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	94	-0.01
20	4,6-Dinitro-2-methylphenol	0.069	0.083	-20.3	116	0.00
21	4-Bromophenyl-phenylether	0.274	0.273	0.4	95	0.00
22	Hexachlorobenzene	0.374	0.371	0.8	98	0.00
23	Atrazine	0.184	0.176	4.3	94	0.00
24	Pentachlorophenol	0.132	0.153	-15.9	121	0.00
25	Phenanthrene	1.167	1.143	2.1	94	-0.01
26	Anthracene	1.062	1.032	2.8	96	0.00
27 SURR	Fluoranthene-d10	0.993	0.942	5.1	90	0.00
28	Fluoranthene	1.361	1.278	6.1	90	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
30	Pyrene	1.624	1.604	1.2	90	0.00
31 S	Terphenyl-d14	0.797	0.773	3.0	86	0.00
32	Benzo(a)anthracene	1.410	1.364	3.3	89	0.00
33	Chrysene	1.475	1.431	3.0	87	0.00
34	Bis(2-ethylhexyl)phthalate	0.574	0.599	-4.4	90	0.00
35 I	Perylene-d12	1.000	1.000	0.0	90	-0.01
36	Indeno(1,2,3-cd)pyrene	1.577	1.479	6.2	94	-0.01
37	Benzo(b)fluoranthene	1.374	1.297	5.6	90	0.00
38	Benzo(k)fluoranthene	1.360	1.318	3.1	94	0.00
39 C	Benzo(a)pyrene	1.190	1.120	5.9	92	0.00
40	Dibenzo(a,h)anthracene	1.255	1.170	6.8	95	-0.02
41	Benzo(g,h,i)perylene	1.403	1.335	4.8	96	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4