

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN011025\
 Data File : BN035912.D
 Acq On : 10 Jan 2025 21:20
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jan 10 22:14:48 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN010225.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jan 02 15:39:17 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	82	0.00
2	1,4-Dioxane	0.400	0.413	-3.2	90	0.00
3	n-Nitrosodimethylamine	0.400	0.439	-9.7	92	-0.01
4 S	2-Fluorophenol	0.400	0.373	6.8	79	0.00
5 S	Phenol-d6	0.400	0.373	6.8	79	0.00
6	bis(2-Chloroethyl)ether	0.400	0.376	6.0	76	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	73	0.00
8 S	Nitrobenzene-d5	0.400	0.444	-11.0	87	-0.01
9	Naphthalene	0.400	0.395	1.3	75	0.00
10	Hexachlorobutadiene	0.400	0.425	-6.2	81	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.385	3.8	72	0.00
12	2-Methylnaphthalene	0.400	0.389	2.8	73	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	67	0.00
14 S	2,4,6-Tribromophenol	0.400	0.459	-14.8	78	-0.01
15 S	2-Fluorobiphenyl	0.400	0.410	-2.5	70	0.00
16	Acenaphthylene	0.400	0.402	-0.5	70	0.00
17	Acenaphthene	0.400	0.418	-4.5	72	0.00
18	Fluorene	0.400	0.394	1.5	68	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	62	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.604	-51.0#	97	0.00
21	4-Bromophenyl-phenylether	0.400	0.429	-7.2	68	0.00
22	Hexachlorobenzene	0.400	0.401	-0.3	65	0.00
23	Atrazine	0.400	0.418	-4.5	68	0.00
24	Pentachlorophenol	0.400	0.470	-17.5	82	0.00
25	Phenanthrene	0.400	0.403	-0.8	64	-0.01
26	Anthracene	0.400	0.397	0.8	65	0.00
27 SURR	Fluoranthene-d10	0.400	0.387	3.3	61	0.00
28	Fluoranthene	0.400	0.382	4.5	61	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	59	0.00
30	Pyrene	0.400	0.406	-1.5	62	0.00
31 S	Terphenyl-d14	0.400	0.384	4.0	57	0.00
32	Benzo(a)anthracene	0.400	0.403	-0.8	62	0.00
33	Chrysene	0.400	0.393	1.8	59	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.477	-19.2	69	0.00
35 I	Perylene-d12	0.400	0.400	0.0	64	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.395	1.3	71	-0.01
37	Benzo(b)fluoranthene	0.400	0.373	6.8	64	0.00
38	Benzo(k)fluoranthene	0.400	0.389	2.8	68	0.00
39 C	Benzo(a)pyrene	0.400	0.384	4.0	68	0.00
40	Dibenzo(a,h)anthracene	0.400	0.395	1.3	72	-0.01
41	Benzo(g,h,i)perylene	0.400	0.388	3.0	70	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4