

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	01/08/25
Project:	RW5B FT WC - # 112G08005-WE13		Date Received:	01/09/25
Client Sample ID:	RW5-SP100-20250108		SDG No.:	Q1040
Lab Sample ID:	Q1040-01		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	990	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035906.D	1	01/10/25 08:10	01/10/25 17:45	PB166005

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	9.60	E	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.42		30 - 150		105%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.47		30 - 150		118%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.46	*	55 - 111		114%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.44	*	53 - 106		111%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.52		58 - 132		131%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1120	7.832				
1146-65-2	Naphthalene-d8	2160	10.622				
15067-26-2	Acenaphthene-d10	1000	14.463				
1517-22-2	Phenanthrene-d10	1940	17.199				
1719-03-5	Chrysene-d12	1720	21.376				
1520-96-3	Perylene-d12	1970	23.681				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products