



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1040

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN035917.D

| Lab Sample ID | Parameter | Spike | Result | Unit | Rec | RPD | Qual | RPD | Low | Limits | |
|---------------|-------------|-------|--------|------|-----|-----|------|------|-----|--------|-----|
| | | | | | | | | Qual | | High | RPD |
| PB166005BS | 1,4-Dioxane | 0.4 | 0.41 | ug/L | 103 | | | | 70 | 130 | |