

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1040

SDG NO.: Q1040

Lab File ID: BN035870.D

DFTPP Injection Date: 01/02/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:49

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	39.9
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	39
70	Less than 2.0% of mass 69	0.3 (0.7) 1
127	10.0 - 80.0% of mass 198	43.4
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.2
275	10.0 - 60.0% of mass 198	25.6
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	9.5
442	Greater than 50% of mass 198	63
443	15.0 - 24.0% of mass 442	12.1 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN035871.D	01/02/2025	11:28
SSTDICC0.2	SSTDICC0.2	BN035872.D	01/02/2025	12:04
SSTDICCC0.4	SSTDICCC0.4	BN035873.D	01/02/2025	12:40
SSTDICC0.8	SSTDICC0.8	BN035874.D	01/02/2025	13:16
SSTDICC1.6	SSTDICC1.6	BN035875.D	01/02/2025	13:52
SSTDICC3.2	SSTDICC3.2	BN035876.D	01/02/2025	14:28
SSTDICC5.0	SSTDICC5.0	BN035877.D	01/02/2025	15:04