

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51921 MA No. : SDG No.: GCPF7
 Initial Calibration Verification Source : MP84065
 Continuing Calibration Verification Source : MP84068
 Run Batch: LB134392 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	1000	1000	101	0	5000	5000	99	0	4900	98	1
Barium	520	550	105	1	10000	10000	104	1	10000	101	2
Cadmium	510	480	95	0	2500	2500	99	0	2500	99	1
Chromium	520	530	102	0	15500	16000	102	0	16000	100	0
Cobalt	520	470	91	0	2500	2500	100	0	2500	99	1
Lead	1000	970	97	0	25000	25000	98	0	24000	98	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51921 MA No. : SDG No.: GCPF7
 Initial Calibration Verification Source : MP84065
 Continuing Calibration Verification Source : MP84068
 Run Batch: LB134392 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV003				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					5000	5000	99	0			
Barium					10000	10000	101	0			
Cadmium					2500	2500	99	0			
Chromium					15500	16000	102	0			
Cobalt					2500	2500	99	0			
Lead					25000	24000	98	0			