

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51921 MA No. : SDG No.: GCPF7
 Analytical Method: ICP-AES ICSA Source: EPA-1211
 Instrument ID: P4 ICSB Source: EPA-0710
 Run Batch: LB134392
 Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	255000	247000	240000	93	230000	92
Arsenic	0	104	-0.67		92	88
Barium	6	537	3.1	51	480	88
Cadmium	1	972	1.5	146	970	100
Calcium	245000	235000	240000	100	230000	100
Chromium	52	542	61	117	600	111
Cobalt	0	476	1.0		490	103
Iron	101000	99300	100000	101	98000	98
Lead	0	49	3.5		53	108
Magnesium	255000	248000	240000	94	230000	92
Potassium	0	0	-90		-470	
Sodium	0	0	-290		-340	