

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51921 MA No.: _____ SDG No.: GCPK5
Instrument ID: P4 Start Date: 01/23/2025
Analytical Method: ICP-AES Run Batch: LB134392
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	0.0	0	10	7.5	-25	1250	1100	-12
Barium	0	0.030	0	200	190	-5	2500	2300	-8
Cadmium	0	-0	0	5	5.1	2	625	630	1
Chromium	0	-0	0	10	11	10	3875	4000	3
Cobalt	0	0.0	0	50	49	-2	625	620	-1
Lead	0	0.0	0	10	8.4	-16	6250	6200	-1

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51921 MA No.: _____ SDG No.: GCPK5
Instrument ID: P4 Start Date: 01/23/2025
Analytical Method: ICP-AES Run Batch: LB134392
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	2500	2400	-4	5000	4700	-6	10000	11000	10
Barium	5000	5000	0	10000	9800	-2	20000	20000	0
Cadmium	1250	1300	4	2500	2500	0	5000	5000	0
Chromium	7750	8200	6	15500	15000	-3	31000	31000	0
Cobalt	1250	1300	4	2500	2400	-4	5000	5000	0
Lead	12500	13000	4	25000	24000	-4	50000	50000	0

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51921 MA No.: SDG No.: GCPK5
Instrument ID: P4 Start Date: 01/23/2025
Analytical Method: ICP-AES Run Batch: LB134392
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	0.0	0						
Barium	0	0.0	0						
Cadmium	0	0.0	0						
Chromium	0	0.0	0						
Cobalt	0	0.0	0						
Lead	0	0.0	0						