



284 Sheffield Street, Mountainside, NJ 07092 Phone: 908 789 8900 Fax: 908 789 8922

CASE NARRATIVE

VERINA CONSULTING GROUP, LLC

Project Name: Rotor Clip NJ WTD - 2025

Project # N/A

Chemtech Project # Q1048

Test Name: Metals Group4

A. Number of Samples and Date of Receipt:

1 Water sample was received on 01/09/2025.

B. Parameters:

According to the Chain of Custody document, the following analyses were requested: Ammonia, Metals Group4 and Residual Chlorine. This data package contains results for Metals Group4.

C. Analytical Techniques:

The analysis of Metals Group4 was based on method 6010D and digestion based on method 3010 (waters).

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Blank Spike met requirements for all samples.

The Duplicate analysis met criteria for all samples.

The Matrix Spike analysis met criteria for all samples.

The Matrix Spike Duplicate analysis met criteria for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Calibration met the requirements.

The Serial Dilution met the acceptable requirements.

E. Additional Comments:

The fax and hardcopy is not matching for Metals Sample Q1048-01 due to at the time of fax, sample analyzed without QC set, but at the time of second review lab noticed QC set was not analyzed therefore this sample analyzed with QC set and reported. Hard copy is reported corrected.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____