



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1048

Project: Rotor Clip - PO# 5183.0001

RunNo.: LB134218

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Residual Chlorine	mg/L	0.413	0.4	103	90-110	01/09/2025
Sample ID: CCV1 Residual Chlorine	mg/L	0.403	0.4	101	90-110	01/09/2025
Sample ID: CCV2 Residual Chlorine	mg/L	0.403	0.4	101	90-110	01/09/2025

Initial and Continuing Calibration Verification

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1048

Project: Rotor Clip - PO# 5183.0001

RunNo.: LB134227

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.97	1	97	90-110	01/10/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.99	1	99	90-110	01/10/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	01/10/2025