

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1048
Project:	Rotor Clip - PO# 5183.0001	Run No.:	LB134218

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134218BS							
Residual Chlorine	mg/L	0.4	0.40		101	1	90-110	01/09/2025

Laboratory Control Sample Summary

Client:	VERINA CONSULTING GROUP, LLC	SDG No.:	Q1048
Project:	Rotor Clip - PO# 5183.0001	Run No.:	LB134227

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165991BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	01/10/2025