



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: VERINA CONSULTING GROUP, LLC

SDG No.: Q1048

Contract: VERI01

Lab Code: CHEM

Case No.: Q1048

SAS No.: Q1048

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Chromium	9.43	10.0	94	40 - 160	P	01/17/2025	11:35	LB134334
	Copper	21.4	20.0	107	40 - 160	P	01/17/2025	11:35	LB134334
	Nickel	38.5	40.0	96	40 - 160	P	01/17/2025	11:35	LB134334
	Zinc	41.2	40.0	103	40 - 160	P	01/17/2025	11:35	LB134334