

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: VERINA CONSULTING GROUP, LLC **SDG No.:** Q1048
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1048 **SAS No.:** Q1048
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Chromium	55.0	52.0	106	42	62	01/17/2025	11:39	LB134334
	Copper	20.6	2.0	1030	-18	22	01/17/2025	11:39	LB134334
	Nickel	2.52	2.0	126	-38	42	01/17/2025	11:39	LB134334
	Zinc	3.10			-40	40	01/17/2025	11:39	LB134334
ICSAB01	Chromium	560	542	103	460	624	01/17/2025	11:43	LB134334
	Copper	496	511	97	434	588	01/17/2025	11:43	LB134334
	Nickel	974	954	102	810	1100	01/17/2025	11:43	LB134334
	Zinc	1050	952	110	809	1095	01/17/2025	11:43	LB134334