

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q1048
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1048 **SAS No.:** Q1048
Matrix: Water **Sample ID:** Q1048-01 **Client ID:** WATER TREATMENT DISCHARGEDUP
Percent Solids for Sample: NA **Duplicate ID** Q1048-01DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	236		231		2		P
Copper	ug/L	20	12.5		12.5		0		P
Nickel	ug/L	20	41.2		40.7		1		P
Zinc	ug/L	20	1180		1160		2		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: VERINA CONSULTING GROUP, LLC **Level:** LOW **SDG No.:** Q1048
Contract: VERI01 **Lab Code:** CHEM **Case No.:** Q1048 **SAS No.:** Q1048
Matrix: Water **Sample ID:** Q1048-01MS **Client ID:** WATER TREATMENT DISCHARGEMSD
Percent Solids for Sample: NA **Duplicate ID** Q1048-01MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Chromium	ug/L	20	421		415		1		P
Copper	ug/L	20	148		146		1		P
Nickel	ug/L	20	286		282		1		P
Zinc	ug/L	20	1230		1220		1		P