
From: Bett, Daisy <Daisy.Bett@gdit.com>
Sent: Tuesday, January 14, 2025 3:20 PM
To: Deepak Parmar; Sohil Jodhani; Mohammad Ahmed
Cc: St-Juste, Reginald; Leung.christina@epa.gov; Feranda, Jennifer; Brandon-Bazile, Kim; Bauer, Heather E; Johnson, Matthew
Subject: RE: Region 02 | Case 51945 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC | FINAL
Attachments: SKM_95825011414110.pdf; CLP CoC 01.10.2024 Updated.pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.


Secured by Check Point

Good afternoon Deepak,

Case 51679 is not the correct Case number. The samples listed on the COC belong to Case 51945. I have re-attached the revised COC for reference.

Thank you,
Daisy Bett
Research Analyst Associate
GDIT Federal Civilian Division
EPA Region 2 CLP QSS Coordinator
Under contract to the EPA

T: 571.454.0186
daisy.bett@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



GENERAL DYNAMICS
Information Technology

Leave alert: none

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not

operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Deepak Parmar <Deepak.Parmar@alliancetg.com>

Sent: Tuesday, January 14, 2025 2:27 PM

To: Bett, Daisy <Daisy.Bett@gdit.com>; Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>; Mohammad Ahmed <mohammad.ahmed@alliancetg.com>

Cc: St-Juste, Reginald <st-juste.reginald@epa.gov>; Leung.christina@epa.gov; Feranda, Jennifer <Feranda.jennifer@epa.gov>; Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Bauer, Heather E <Heather.Bauer@gdit.com>; Johnson, Matthew <Matthew.Johnson32@gdit.com>

Subject: RE: Region 02 | Case 51945 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC | FINAL

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good afternoon,

Case number 51679 on COC is not correct case number ?

Thanks & Regards,



Deepak Parmar

QA/QC

An Alliance Technical Group Company

Main: 908-789-8900

Direct: 908-728-3154

Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com



From: Bett, Daisy <Daisy.Bett@gdit.com>

Sent: Tuesday, January 14, 2025 12:52 PM

To: Deepak Parmar <Deepak.Parmar@AllianceTG.com>; Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>; Mohammad Ahmed <Mohammad.Ahmed@AllianceTG.com>

Cc: St-Juste, Reginald <st-juste.reginald@epa.gov>; Leung.christina@epa.gov; Feranda, Jennifer <Feranda.jennifer@epa.gov>; Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Bauer, Heather E <Heather.Bauer@gdit.com>; Johnson, Matthew <Matthew.Johnson32@gdit.com>

Subject: Region 02 | Case 51945 | Lab ACE | Issue Discrepancies with tags, jars, and/or COC | FINAL

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon,


Issue: The COC shipped with samples for Case 51945 on 01/10/2025 listed an incorrect Case number.

Resolution: Per Region 2, a revised COC with the correct Case number listed is attached. Please note the issue in the SDG Narrative and proceed with the analysis of the samples.

Please note that the laboratory may contact the appropriate CLP PM should any defects need to be waived for this issue.

Thank you,
Daisy Bett
Research Analyst Associate
GDIT Federal Civilian Division
EPA Region 2 CLP QSS Coordinator
Under contract to the EPA

T: 571.454.0186
daisy.bett@gdit.com
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com



GENERAL DYNAMICS
DEFENSE TECHNOLOGY

Leave alert: none

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: St-Juste, Reginald <st-juste.reginald@epa.gov>
Sent: Tuesday, January 14, 2025 11:18 AM
To: Bett, Daisy <Daisy.Bett@gdit.com>
Cc: Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>; Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Feranda, Jennifer <Feranda.Jennifer@epa.gov>
Subject: FW: Updated chain for CLP Case # 51945

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hi Daisy,

Looks like the Case # provided on the chain given to the CLP had the incorrect Case #.

The updated chain with the correct Case # 51945 is attached.

See e-mail below.

Thank you,

Reginald St-Juste
Hazardous Waste Support Section
US EPA Region 2
2890 Woodbridge Ave.
Edison, NJ 08837
O: 732-321-6746

From: Robinson, Noah <nrobinson@eaest.com>
Sent: Tuesday, January 14, 2025 11:07 AM
To: St-Juste, Reginald <st-juste.reginald@epa.gov>
Cc: Casey, Robert <rcasey@eaest.com>; Tara Lamond <tlamond@eaest.com>; Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>
Subject: FW: Updated chain for CLP Case # 51945

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Hi Regi,

I received an out of office response from Christina.

Please see below.

Thank you,

Noah Robinson
Scientist II



EA Science and Technology

Office:

8055 E Tufts Ave. Suite 220

Denver, Colorado 80237

Office: 315.565.6552 (Direct Line)

Mobile: 207.400.1356

nrobinson@eaest.com

From: Robinson, Noah
Sent: Tuesday, January 14, 2025 9:03 AM
To: Leung.Christina@epa.gov
Cc: Casey, Robert <rcasey@eaest.com>; Lamond, Tara <tlamond@eaest.com>
Subject: Updated chain for CLP Case # 51945

Hi Christina,

We noticed during review that the Case # provided on the chain given to the CLP had the incorrect Case #.

We have attached an updated chain with the correct Case # 51945.

The trip report will list the correct case number as well once compiled.

Let me know if there are any questions.

Thank you,

Noah Robinson

Scientist II



EA Science and Technology

Office:

8055 E Tufts Ave. Suite 220

Denver, Colorado 80237

Office: 315.565.6552 (Direct Line)

Mobile: 207.400.1356

nrobinson@eaest.com

From: Panday, Prakriti P <Prakriti.Panday@gdit.com>
Sent: Thursday, January 23, 2025 2:44 PM
Subject: TASK AREA DAS | REGION 2 | CASE 51945 | LAB ACE | SDG BH8T0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL
Attachments: Q1057-07(BH8T6) 10x.pdf; Q1057-09(BH8T8) 10x.pdf; Q1057-10(BH8T9) 10x.pdf; Q1057-16(BH8W5) 10x.pdf; Q1057-03-07-9-10-16.jpg; Q1057-14(BH8W3).pdf; Q1057-20(BH8W9).pdf; Q1057-03(BH8T2).pdf; Q1057-07(BH8T6).pdf; Q1057-09(BH8T8).pdf; Q1057-10(BH8T9).pdf; Q1057-16(BH8W5).pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon,

Please refer to the Region's response below regarding Case 51945/SDG BH8T0. Also, please be sure to provide detailed information about this issue in the SDG Narrative and let me know if you have any questions.

ISSUE: " 'The Lab has received water samples for TVOA analysis. The samples BH8T2, BH8T6, BH8T8, BH8T9, BH8W5 were received with foamy nature of the samples and samples could not be analyzed undiluted due to foamy matrix. Therefore, as a precautionary step, Lab has analyzed samples BH8T2, BH8T6, BH8T8, BH8T9 and BH8W5 with most plausible dilution factors. Please see attached quant reports and samples image for your reference. In this case, lab would like to confirm that lab will report all samples diluted analysis as first analysis in the electronic deliverables.' "

The lab has analyzed samples BH8W3 and BH8W9 in a continuous analytical sequence for TVOA analysis. Both samples required dilution to bring target analytes within calibration range. Therefore, instrument blank was not analyzed in between the samples due to continuous analytical sequence. In this case, the lab would like to confirm that the lab will report undiluted TVOA analysis without instrument blank between the samples and further dilution analysis in final electronic deliverables.

REGION: "Could you please provide the dilution factors used. Thanks."

LAB: "All provided a quant report have dilution factors. The diluting factor was not included in the subject line for the attachments. Please find attachment with dilution factor in subject line for the sample BH8T2, BH8T6, BH8T8, BH8T9, BH8W5."

REGION: "Please advise the laboratory to proceed and state their rational for any deviations from SOW in the SDG narrative. Going forward, please request the laboratory to state the status of laboratory QC in dilution runs."

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

PRAKRITI PANDAY
Associate Environmental Analyst
571-533-3261

GDIT Federal Civilian Division
Under Contract to EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Kumar, Narendra <Kumar.Narendra@epa.gov>

Sent: Thursday, January 23, 2025 2:32 PM

To: Panday, Prakriti P <Prakriti.Panday@gdit.com>

Cc: Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Arnone, Russell <Arnone.Russell@epa.gov>; Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>; Feranda, Jennifer <Feranda.Jennifer@epa.gov>; Capil, Shobitha <Capil.Shobitha@epa.gov>; McHugh, Lauren (she/her/hers) <mchugh.lauren@epa.gov>

Subject: RE: TASK AREA DAS | REGION 2 | CASE 51945 | LAB ACE | SDG BH8T0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello Prakriti,
Please advise the laboratory to proceed and state their rationale for any deviations from SOW in the SDG narrative. Going forward, please request the laboratory to state the status of laboratory QC in dilution runs.
Thanks.

Narendra Kumar

From: Panday, Prakriti P <Prakriti.Panday@gdit.com>

Sent: Thursday, January 23, 2025 11:05 AM

To: Kumar, Narendra <Kumar.Narendra@epa.gov>; Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Arnone, Russell <Arnone.Russell@epa.gov>; Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>; Feranda, Jennifer <Feranda.Jennifer@epa.gov>; Capil, Shobitha <Capil.Shobitha@epa.gov>; McHugh, Lauren (she/her/hers) <mchugh.lauren@epa.gov>

Subject: TASK AREA DAS | REGION 2 | CASE 51945 | LAB ACE | SDG BH8T0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good morning,

Please see the laboratory's response for Case 51945/SDG BH8T0 and please advise.

ISSUE: “ ‘The Lab has received water samples for TVOA analysis. The samples BH8T2, BH8T6, BH8T8, BH8T9, BH8W5 were received with foamy nature of the samples and samples could not be analyzed undiluted due to foamy matrix. Therefore, as a precautionary step, Lab has analyzed samples BH8T2, BH8T6, BH8T8, BH8T9 and BH8W5 with **most plausible dilution factors**. Please see attached quant reports and samples image for your reference. In this case, lab would like to confirm that lab will report all samples diluted analysis as first analysis in the electronic deliverables.’

The lab has analyzed samples BH8W3 and BH8W9 in a continuous analytical sequence for TVOA analysis. Both samples required dilution to bring target analytes within calibration range. Therefore, instrument blank was not analyzed in between the samples due to continuous analytical sequence. In this case, the lab would like to confirm that the lab will report undiluted TVOA analysis without instrument blank between the samples and further dilution analysis in final electronic deliverables.

REGION: “Could you please provide the dilution factors used. Thanks.”

LAB: “All provided a quant report have dilution factors. The diluting factor was not included in the subject line for the attachments. Please find attachment with dilution factor in subject line for the sample BH8T2, BH8T6, BH8T8, BH8T9, BH8W5.”

Thank you,

PRAKRITI PANDAY
Associate Environmental Analyst
571-533-3261
GDIT Federal Civilian Division
Under Contract to EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Nimisha Pandya <Nimisha.Pandya@alliancetg.com>

Sent: Thursday, January 23, 2025 10:23 AM

To: Panday, Prakriti P <Prakriti.Panday@gdit.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>

Subject: Re: TASK AREA DAS | REGION 2 | CASE 51945 | LAB ACE | SDG BH8T0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good morning, Prakrit,

All provided a quant report have dilution factors. The diluting factor was not included in the subject line for the attachments. Please find attachment with dilution factor in subject line for the sample BH8T2, BH8T6, BH8T8, BH8T9, BH8W5.

Thanks & Regards,

Nimisha N Pandya
QA/QC supervisor
An Alliance Technical Group Company
Main: 908-789-8900

Direct: 908-728-3153
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com     

From: Panday, Prakriti P <Prakriti.Panday@gdit.com>
Sent: Thursday, January 23, 2025 9:39 AM
To: Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>
Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Sohil Jodhani <Sohil.Jodhani@AllianceTG.com>
Subject: TASK AREA DAS | REGION 2 | CASE 51945 | LAB ACE | SDG BH8T0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good morning,

Please provide a response to the Region's follow up question for Case 51945/SDG BH8T0.

ISSUE: " 'The Lab has received water samples for TVOA analysis. The samples BH8T2, BH8T6, BH8T8, BH8T9, BH8W5 were received with foamy nature of the samples and samples could not be analyzed undiluted due to foamy matrix. Therefore, as a precautionary step, Lab has analyzed samples BH8T2, BH8T6, BH8T8, BH8T9 and BH8W5 with **most plausible dilution factors**. Please see attached quant reports and samples image for your reference. In this case, lab would like to confirm that lab will report all samples diluted analysis as first analysis in the electronic deliverables.'

The lab has analyzed samples BH8W3 and BH8W9 in a continuous analytical sequence for TVOA analysis. Both samples required dilution to bring target analytes within calibration range. Therefore, instrument blank was not analyzed in between the samples due to continuous analytical sequence. In this case, the lab would like to confirm that the lab will report undiluted TVOA analysis without instrument blank between the samples and further dilution analysis in final electronic deliverables.

REGION: "Could you please provide the dilution factors used. Thanks."

Thank you,

PRAKRITI PANDAY

Associate Environmental Analyst

571-533-3261

GDIT Federal Civilian Division

Under Contract to EPA

15036 Conference Center Drive

Chantilly, VA 20151

www.gdit.com

GENERAL DYNAMICS
Information Technology

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Kumar, Narendra <Kumar.Narendra@epa.gov>

Sent: Thursday, January 23, 2025 9:21 AM

To: Panday, Prakriti P <Prakriti.Panday@gdit.com>

Cc: Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Arnone, Russell <Arnone.Russell@epa.gov>; Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>; Feranda, Jennifer <Feranda.Jennifer@epa.gov>; Capil, Shobitha <Capil.Shobitha@epa.gov>; McHugh, Lauren (she/her/hers) <mchugh.lauren@epa.gov>

Subject: RE: TASK AREA DAS | REGION 2 | CASE 51945 | LAB ACE | SDG BH8T0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello Prakriti,

Could you please provide the dilution factors used. Thanks.

Narendra Kumar

From: Panday, Prakriti P <Prakriti.Panday@gdit.com>

Sent: Wednesday, January 22, 2025 5:01 PM

To: Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Arnone, Russell <Arnone.Russell@epa.gov>; Kumar, Narendra <Kumar.Narendra@epa.gov>; Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>; Feranda, Jennifer <Feranda.Jennifer@epa.gov>

Subject: TASK AREA DAS | REGION 2 | CASE 51945 | LAB ACE | SDG BH8T0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good afternoon,

Please advise the laboratory with Case 51945/SDG BH8T0.

ISSUE: “ ‘The Lab has received water samples for TVOA analysis. The samples BH8T2, BH8T6, BH8T8, BH8T9, BH8W5 were received with foamy nature of the samples and samples could not be analyzed undiluted due to foamy matrix. Therefore, as a precautionary step, Lab has analyzed samples BH8T2, BH8T6, BH8T8, BH8T9 and BH8W5 with **most plausible dilution factors**. Please see attached quant reports and samples image for your reference. In this case, lab would like to confirm that lab will report all samples diluted analysis as first analysis in the electronic deliverables.’

The lab has analyzed samples BH8W3 and BH8W9 in a continuous analytical sequence for TVOA analysis. Both samples required dilution to bring target analytes within calibration range. Therefore, instrument blank was not analyzed in between the samples due to continuous analytical sequence. In this case, the lab would like to confirm that the lab will report undiluted TVOA analysis without instrument blank between the samples and further dilution analysis in final electronic deliverables.

Please see attached.”

Thank you,

PRAKRITI PANDAY

Associate Environmental Analyst

571-533-3261

GDIT Federal Civilian Division

Under Contract to EPA

15036 Conference Center Drive

Chantilly, VA 20151

www.gdit.com

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: DASsupport

Sent: Wednesday, January 22, 2025 3:56 PM

To: 'Nimisha Pandya' <Nimisha.Pandya@alliancetg.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Sohil Jodhani <Sohil.Jodhani@alliancetg.com>

Subject: TASK AREA DAS | REGION 2 | CASE 51945 | LAB ACE | SDG BH8T0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Thank you for contacting DAS Support. This is to confirm that your email has been received and is currently being addressed. An updated response will be provided as soon as it is available.

Thank you,

The DAS Support Team

DASsupport@gdit.com

GDIT Federal Civilian Division

Under Contract to EPA

15036 Conference Center Drive

Chantilly, VA 20151

www.gdit.com

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Nimisha Pandya <Nimisha.Pandya@alliancetg.com>

Sent: Wednesday, January 22, 2025 3:47 PM

To: DASsupport <DASsupport@gdit.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Sohil Jodhani <Sohil.Jodhani@alliancetg.com>

Subject: REGION 2 | CASE 51945 | LAB ACE | SDG BH8T0 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Good afternoon,

The lab is sending this email regarding Case 51945 and SDG BH8T0.

‘The Lab has received water samples for TVOA analysis. The samples BH8T2, BH8T6, BH8T8, BH8T9, BH8W5 were received with foamy nature of the samples and samples could not be analyzed undiluted due to foamy matrix. Therefore, as a precautionary step, Lab has analyzed samples BH8T2, BH8T6, BH8T8, BH8T9 and BH8W5 with most plausible dilution factors. Please see attached quant reports and samples image for your reference. In this case, lab would like to confirm that lab will report all samples diluted analysis as first analysis in the electronic deliverables.’

The lab has analyzed samples BH8W3 and BH8W9 in a continuous analytical sequence for TVOA analysis. Both samples required dilution to bring target analytes within calibration range. Therefore, instrument blank was not analyzed in between the samples due to continuous analytical sequence. In this case, the lab would like

to confirm that the lab will report undiluted TVOA analysis without instrument blank between the samples and further dilution analysis in final electronic deliverables.

Please see attached.

Thanks & Regards,

Nimisha N Pandya
QA/QC supervisor
An Alliance Technical Group Company
Main: 908-789-8900

Direct: 908-728-3153
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092

www.alliancetg.com     