

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1060 SAS No.: Q1060 SDG NO.: Q1060

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/20/2024 12/20/2024

Client Sample No.: CCAL01 Date Analyzed: 01/14/2025

Lab Sample No.: AR1660CCC400 Data File : PR069964.D Time Analyzed: 11:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.851	4.751	4.951	30435500.000	27986100.000	-8.0
Aroclor-1016-2	4.872	4.772	4.972	46509090.000	44285200.000	-4.8
Aroclor-1016-3	4.936	4.836	5.036	28773090.000	27528600.000	-4.3
Aroclor-1016-4	5.038	4.939	5.139	22386480.000	20801800.000	-7.1
Aroclor-1016-5	5.353	5.254	5.454	22923650.000	22766300.000	-0.7
Aroclor-1260-1	6.565	6.466	6.666	40163220.000	40122300.000	-0.1
Aroclor-1260-2	6.848	6.748	6.948	44680180.000	43224600.000	-3.3
Aroclor-1260-3	7.239	7.138	7.338	35412650.000	35316800.000	-0.3
Aroclor-1260-4	7.482	7.381	7.581	39008850.000	38948400.000	-0.2
Aroclor-1260-5	7.821	7.722	7.922	70337060.000	66522100.000	-5.4
DCB	9.519	9.421	9.621	515829000.000	528087000.000	2.4
TCX	3.629	3.530	3.730	1039115000.000	1049760000.000	1.0

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1060 SAS No.: Q1060 SDG NO.: Q1060

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/20/2024 12/20/2024

Client Sample No.: CCAL01 Date Analyzed: 01/14/2025

Lab Sample No.: AR1660CCC400 Data File : PR069964.D Time Analyzed: 11:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.078	3.982	4.182	175384700.000	168470000.000	-3.9
Aroclor-1016-2	4.097	4.000	4.200	257002400.000	248713000.000	-3.2
Aroclor-1016-3	4.276	4.180	4.380	133980000.000	131896000.000	-1.6
Aroclor-1016-4	4.327	4.231	4.431	109835400.000	114464000.000	4.2
Aroclor-1016-5	4.546	4.450	4.650	134578200.000	139390000.000	3.6
Aroclor-1260-1	5.641	5.546	5.746	235212500.000	227845000.000	-3.1
Aroclor-1260-2	5.848	5.753	5.953	268390900.000	260948000.000	-2.8
Aroclor-1260-3	6.006	5.911	6.111	255470300.000	253596000.000	-0.7
Aroclor-1260-4	6.513	6.417	6.617	217227400.000	213831000.000	-1.6
Aroclor-1260-5	6.780	6.684	6.884	499565500.000	472444000.000	-5.4
DCB	8.232	8.137	8.337	3355869000.000	3342410000.000	-0.4
TCX	2.960	2.862	3.062	5613354000.000	5825180000.000	3.8