

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1060 SAS No.: Q1060 SDG NO.: Q1060

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/20/2024 12/20/2024

Client Sample No.: CCAL02 Date Analyzed: 01/14/2025

Lab Sample No.: AR1660CCC400 Data File : PR069968.D Time Analyzed: 12:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.848	4.751	4.951	30435500.000	27362700.000	-10.1
Aroclor-1016-2	4.870	4.772	4.972	46509090.000	43977400.000	-5.4
Aroclor-1016-3	4.934	4.836	5.036	28773090.000	27030600.000	-6.1
Aroclor-1016-4	5.036	4.939	5.139	22386480.000	20727500.000	-7.4
Aroclor-1016-5	5.351	5.254	5.454	22923650.000	22472400.000	-2.0
Aroclor-1260-1	6.563	6.466	6.666	40163220.000	41008000.000	2.1
Aroclor-1260-2	6.846	6.748	6.948	44680180.000	43884900.000	-1.8
Aroclor-1260-3	7.236	7.138	7.338	35412650.000	35592200.000	0.5
Aroclor-1260-4	7.479	7.381	7.581	39008850.000	38637500.000	-1.0
Aroclor-1260-5	7.818	7.722	7.922	70337060.000	64436200.000	-8.4
DCB	9.515	9.421	9.621	515829000.000	528670000.000	2.5
TCX	3.628	3.530	3.730	1039115000.000	1043670000.000	0.5

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1060 SAS No.: Q1060 SDG NO.: Q1060

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/20/2024 12/20/2024

Client Sample No.: CCAL02 Date Analyzed: 01/14/2025

Lab Sample No.: AR1660CCC400 Data File : PR069968.D Time Analyzed: 12:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.078	3.982	4.182	175384700.000	168164000.000	-4.1
Aroclor-1016-2	4.096	4.000	4.200	257002400.000	249880000.000	-2.8
Aroclor-1016-3	4.275	4.180	4.380	133980000.000	131270000.000	-2.0
Aroclor-1016-4	4.326	4.231	4.431	109835400.000	109566000.000	-0.2
Aroclor-1016-5	4.545	4.450	4.650	134578200.000	140044000.000	4.1
Aroclor-1260-1	5.640	5.546	5.746	235212500.000	232074000.000	-1.3
Aroclor-1260-2	5.847	5.753	5.953	268390900.000	264175000.000	-1.6
Aroclor-1260-3	6.005	5.911	6.111	255470300.000	256571000.000	0.4
Aroclor-1260-4	6.512	6.417	6.617	217227400.000	216692000.000	-0.2
Aroclor-1260-5	6.778	6.684	6.884	499565500.000	474078000.000	-5.1
DCB	8.229	8.137	8.337	3355869000.000	3412820000.000	1.7
TCX	2.960	2.862	3.062	5613354000.000	5785870000.000	3.1