

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122024\
 Data File : PR069800.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 20 Dec 2024 18:59
 Operator : AJ\MA
 Sample : AR1262ICC100
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12621214

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 12/23/2024
 Supervised By :mohammad ahmed 12/25/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Dec 21 02:25:55 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR122024CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Sat Dec 21 02:25:02 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.630	2.961	5244520	29128665	5.043	4.915
2) SA Decachlor...	9.520	8.239	5610628	36990237	10.542	10.533
Target Compounds						
36) L8 AR-1262-1	7.240	6.244	5247172	34207668	99.835	107.107
37) L8 AR-1262-2	7.822	6.785	8433504	57524355	105.687	103.017
38) L8 AR-1262-3	8.129	7.090	5771645	23387423	105.038m	104.078
39) L8 AR-1262-4	8.213	7.158	4515307	43593746	104.703	103.154
40) L8 AR-1262-5	8.835	7.697	2679694	19542562	101.890	102.265

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR122024\
Data File : PR069800.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 20 Dec 2024 18:59
Operator : AJ\MA
Sample : AR1262ICC100
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12621214

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 12/23/2024
Supervised By :mohammad ahmed 12/25/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Dec 21 02:25:55 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR122024CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Sat Dec 21 02:25:02 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

