

Surrogate Summary

SDG No.: Q1060

Client: Chemtech Consulting Group

Analytical Method: SFAM_PCB

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PR069769.D	PIBLK-PR069769.D	TCX	1	20	0.20	98		30	150
		DCB	1	40	0.40	100		30	150
		TCX	2	20	0.21	104		30	150
		DCB	2	40	0.40	100		30	150
I.BLK-PR069963.D	PIBLK-PR069963.D	TCX	1	20	0.22	108		30	150
		DCB	1	40	0.45	113		30	150
		TCX	2	20	0.24	121		30	150
		DCB	2	40	0.45	114		30	150
Q1060-13	PCB-GPC2-BLANK	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
Q1060-14	PCB-GPC2-BLANK-SPIKE	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
I.BLK-PR069967.D	PIBLK-PR069967.D	TCX	1	20	0.22	110		30	150
		DCB	1	40	0.45	112		30	150
		TCX	2	20	0.24	120		30	150
		DCB	2	40	0.46	115		30	150