

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1060 SAS No.: Q1060 SDG NO.: Q1060

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 12/19/2024 12/19/2024

Client Sample No.: CCAL01 Date Analyzed: 01/14/2025

Lab Sample No.: INDA348 Data File : PD087390.D Time Analyzed: 11:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.714	6.644	6.784	2592290.000	2663890.000	2.8
4,4'-DDT	7.030	6.960	7.100	2680940.000	2580410.000	-3.7
alpha-BHC	4.006	3.956	4.056	4034650.000	4107680.000	1.8
DCB	9.087	8.987	9.187	3054180.000	3128260.000	2.4
Dieldrin	6.356	6.286	6.426	3616610.000	3713190.000	2.7
Endosulfan I	6.083	6.013	6.153	3340570.000	3406510.000	2.0
Endrin	6.583	6.513	6.653	2987460.000	2964960.000	-0.8
gamma-BHC (Lindane)	4.337	4.287	4.387	3905620.000	3971440.000	1.7
Heptachlor	4.938	4.888	4.988	3968440.000	4008680.000	1.0
Methoxychlor	7.502	7.432	7.572	1487720.000	1405530.000	-5.5
TCX	3.557	3.507	3.607	2047380.000	2075310.000	1.4

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1060 SAS No.: Q1060 SDG NO.: Q1060

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 12/19/2024 12/19/2024

Client Sample No.: CCAL01 Date Analyzed: 01/14/2025

Lab Sample No.: INDA348 Data File : PD087390.D Time Analyzed: 11:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	5.937	5.867	6.007	12176700.000	12967500.000	6.5
4,4'-DDT	6.192	6.122	6.262	12607700.000	12841800.000	1.9
alpha-BHC	3.397	3.347	3.447	18228000.000	19848100.000	8.9
DCB	8.084	7.984	8.184	13172900.000	14146800.000	7.4
Dieldrin	5.520	5.450	5.590	15309000.000	16396600.000	7.1
Endosulfan I	5.254	5.184	5.324	14494200.000	15504600.000	7.0
Endrin	5.796	5.726	5.866	13970900.000	14172100.000	1.4
gamma-BHC (Lindane)	3.734	3.684	3.784	16919200.000	18276900.000	8.0
Heptachlor	4.088	4.038	4.138	17290400.000	18148100.000	5.0
Methoxychlor	6.763	6.693	6.833	5830000.000	5299760.000	-9.1
TCX	2.883	2.833	2.933	11636900.000	12708900.000	9.2